

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050222\
 Data File : VW022748.D
 Acq On : 02 May 2022 18:40
 Operator : SY/VA
 Sample : N2639-08
 Misc : 4.88g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EXL18

Quant Time: May 03 01:45:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue May 03 01:41:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	322169	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	280627	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	101851	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	165889	25.098	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	100.400%
7) Chloroethane-d5	2.886	69	104776	24.365	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.480%
11) 1,1-Dichloroethene-d2	4.013	63	156082	18.275	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.120%
21) 2-Butanone-d5	7.080	46	56201	44.720	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	89.440%
24) Chloroform-d	7.647	84	238535	25.272	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	101.080%
26) 1,2-Dichloroethane-d4	8.305	65	141358	27.306	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	109.240%
32) Benzene-d6	8.275	84	462925	28.620	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	114.480%
36) 1,2-Dichloropropane-d6	9.275	67	136000	29.932	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	119.720%
41) Toluene-d8	10.323	98	419131	26.824	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	107.280%
43) trans-1,3-Dichloroprop...	10.573	79	51386	24.562	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	98.240%
47) 2-Hexanone-d5	10.921	63	42351	48.827	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.660%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	115600	27.296	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	109.200%
66) 1,2-Dichlorobenzene-d4	13.853	152	92800	26.065	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	104.240%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050222\
Data File : VW022748.D
Acq On : 02 May 2022 18:40
Operator : SY/VA
Sample : N2639-08
Misc : 4.88g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EXL18

Quant Time: May 03 01:45:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Tue May 03 01:41:22 2022
Response via : Initial Calibration

