

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050324\
 Data File : VW028584.D
 Acq On : 03 May 2024 11:19
 Operator : SY/MD
 Sample : P2349-12
 Misc : 5.05g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK002

Quant Time: May 03 23:00:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050224SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri May 03 23:00:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	272770	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	257044	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	115487	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	120589	23.281	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.120%
7) Chloroethane-d5	2.899	69	89881	22.662	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.640%
11) 1,1-Dichloroethene-d2	4.027	65	49395	22.825	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	91.280%
21) 2-Butanone-d5	7.081	46	59453	43.995	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.980%
24) Chloroform-d	7.648	84	199253	23.378	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.520%
26) 1,2-Dichloroethane-d4	8.307	65	116729	23.922	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	95.680%
32) Benzene-d6	8.276	84	396747	23.134	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.520%
36) 1,2-Dichloropropane-d6	9.270	67	118882	23.037	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.160%
41) Toluene-d8	10.319	98	350582	23.084	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.320%
43) trans-1,3-Dichloroprop...	10.575	79	46482	22.390	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.560%
47) 2-Hexanone-d5	10.922	63	39543	44.427	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.860%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	96647	22.334	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.320%
66) 1,2-Dichlorobenzene-d4	13.849	152	104721	23.791	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050324\
Data File : VW028584.D
Acq On : 03 May 2024 11:19
Operator : SY/MD
Sample : P2349-12
Misc : 5.05g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK002

Quant Time: May 03 23:00:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050224SMA.M
Quant Title : SFAM01.0
QLast Update : Fri May 03 23:00:08 2024
Response via : Initial Calibration

