




Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW050719\
Data File : VW010307.D
Acq On : 07 May 2019 19:27
Operator : SY/VA
Sample : VIBLK
Misc : 5.00G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 08 06:53:19 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM050719S.M
Quant Title : VOC Analysis
QLast Update : Wed May 08 04:42:42 2019
Response via : Initial Calibration



TIC: VW010307.D

(16) Methylene chloride (T)

4.915min (+0.000) 1.08ug/L

response 5775

Ion	Exp%	Act%
84.00	100	100
49.00	134.90	122.19
86.00	62.60	53.13
0.00	0.00	0.00