




Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW050719\
Data File : VW010314.D
Acq On : 07 May 2019 22:46
Operator : SY/VA
Sample : K2613-22
Misc : 5.59G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 08 04:40:00 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM050719S.M
Quant Title : VOC Analysis
QLast Update : Wed May 08 02:05:40 2019
Response via : Initial Calibration



TIC: VW010314.D

(16) Methylene chloride (T)

4.909min (-0.006) 2.14ug/L m

response 16023

Ion	Exp%	Act%
84.00	100	100
49.00	134.90	132.29
86.00	62.60	64.75
0.00	0.00	0.00