

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW050719\
 Data File : VW010308.D
 Acq On : 07 May 2019 19:55
 Operator : SY/VA
 Sample : K2613-16
 Misc : 4.86G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFHF7

Quant Time: May 08 04:39:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM050719S.M
 Quant Title : VOC Analysis
 QLast Update : Wed May 08 02:05:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	499531	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	482937	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	197449	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	113032	19.43	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.72%
7) Chloroethane-d5	2.89	69	92228	18.95	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.80%
10) 1,1-Dichloroethene-d2	4.01	63	163421	11.28	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	45.12%
20) 2-Butanone-d5	7.07	46	137885	51.17	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.34%
24) Chloroform-d	7.64	84	340753	25.61	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	102.44%
26) 1,2-Dichloroethane-d4	8.30	65	204203	25.92	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	103.68%
29) Benzene-d6	8.27	84	643184	24.52	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	98.08%
33) 1,2-Dichloropropane-d6	9.27	67	199729	24.91	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	99.64%
37) Toluene-d8	10.32	98	576082	24.63	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	98.52%
38) trans-1,3-Dichloropropene-	10.58	79	89806	24.49	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	97.96%
39) 2-Hexanone-d5	10.93	63	106595	52.02	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	104.04%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	149503	19.60	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	78.40%
61) 1,2-Dichlorobenzene-d4	13.85	152	182186	27.04	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	108.16%


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
13) Acetone	4.12	43	3427	1.486	ug/L	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW050719\
Data File : VW010308.D
Acq On : 07 May 2019 19:55
Operator : SY/VA
Sample : K2613-16
Misc : 4.86G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFHF7

Quant Time: May 08 04:39:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM050719S.M
Quant Title : VOC Analysis
QLast Update : Wed May 08 02:05:40 2019
Response via : Initial Calibration





#13
 Acetone
 Concen: 1.486 ug/L
 RT: 4.12 min Scan# 364
 Delta R.T. 0.01 min
 Lab File: VW010308.D
 Acq: 07 May 2019 19:55

Instrument :
 MSVOA_W
 ClientSampleId :
 BFHF7

Tgt Ion: 43 Resp: 3427
 Ion Ratio Lower Upper
 43 100
 58 23.3 0.0 55.2

