

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051022\
 Data File : VW022948.D
 Acq On : 10 May 2022 11:53
 Operator : SY/VA
 Sample : N2713-08RE
 Misc : 5.78g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBSK8

Quant Time: May 11 01:21:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed May 11 01:19:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	358289	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	351166	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	176702	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	105557	14.360	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	57.440%
7) Chloroethane-d5	2.885	69	86163	18.017	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.080%
11) 1,1-Dichloroethene-d2	4.019	63	115915	12.204	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	48.800%
21) 2-Butanone-d5	7.080	46	62668	44.839	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	89.680%
24) Chloroform-d	7.653	84	206656	19.687	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	78.760%
26) 1,2-Dichloroethane-d4	8.421	65	180	0.031	ug/L	0.12
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.120%#
32) Benzene-d6	8.275	84	382002	18.873	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	75.480%
36) 1,2-Dichloropropane-d6	9.275	67	116648	20.516	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	82.080%
41) Toluene-d8	10.323	98	353530	18.081	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	72.320%
43) trans-1,3-Dichloroprop...	10.579	79	47438	18.120	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	72.480%
47) 2-Hexanone-d5	10.927	63	51232	47.201	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.400%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	117648	22.200	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	88.800%
66) 1,2-Dichlorobenzene-d4	13.853	152	132876	21.512	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	86.040%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051022\
Data File : VW022948.D
Acq On : 10 May 2022 11:53
Operator : SY/VA
Sample : N2713-08RE
Misc : 5.78g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBSK8

Quant Time: May 11 01:21:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Wed May 11 01:19:10 2022
Response via : Initial Calibration

