

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051022\
 Data File : VW022955.D
 Acq On : 10 May 2022 14:42
 Operator : SY/VA
 Sample : N2746-06
 Misc : 5.61g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBSL2

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/11/2022
 Supervised By :Mahesh Dadoda 05/16/2022

Quant Time: May 11 01:22:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed May 11 01:19:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	3745	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	2469m	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	114	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.367	65	1422	18.507	ug/L	0.01
Spiked Amount 25.000	Range 30 - 150		Recovery =	74.040%		
7) Chloroethane-d5	2.904	69	1213	24.266	ug/L	0.02
Spiked Amount 25.000	Range 30 - 150		Recovery =	97.080%		
11) 1,1-Dichloroethene-d2	4.037	63	2099m	21.143	ug/L	0.02
Spiked Amount 25.000	Range 45 - 110		Recovery =	84.560%		
21) 2-Butanone-d5	0.000	46	0d	0.000	ug/L	
Spiked Amount 50.000	Range 20 - 135		Recovery =	0.000%#		
24) Chloroform-d	7.647	84	2881m	26.258	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	105.040%		
26) 1,2-Dichloroethane-d4	8.317	65	1358m	22.567	ug/L	0.01
Spiked Amount 25.000	Range 70 - 130		Recovery =	90.280%		
32) Benzene-d6	8.281	84	4555	32.007	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	128.040%		
36) 1,2-Dichloropropane-d6	9.274	67	1213m	30.343	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	121.360%#		
41) Toluene-d8	10.323	98	3091	22.484	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	89.920%		
43) trans-1,3-Dichloroprop...	0.000	79	0d	0.000	ug/L	
Spiked Amount 25.000	Range 30 - 135		Recovery =	0.000%#		
47) 2-Hexanone-d5	0.000	63	0d	0.000	ug/L	
Spiked Amount 50.000	Range 20 - 135		Recovery =	0.000%#		
56) 1,1,2,2-Tetrachloroeth...	0.000	84	0d	0.000	ug/L	
Spiked Amount 25.000	Range 45 - 120		Recovery =	0.000%#		
66) 1,2-Dichlorobenzene-d4	13.853	152	268	67.251	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	269.000%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051022\
Data File : VW022955.D
Acq On : 10 May 2022 14:42
Operator : SY/VA
Sample : N2746-06
Misc : 5.61g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBSL2

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 05/11/2022
Supervised By : Mahesh Dadoda 05/16/2022

Quant Time: May 11 01:22:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Wed May 11 01:19:10 2022
Response via : Initial Calibration

