

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW051119\
 Data File : VW010391.D
 Acq On : 11 May 2019 19:57
 Operator : SY/VA
 Sample : K2719-04
 Misc : 4.08G/5ML/MSVOA W/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 P026-SS008-0206-02

Quant Time: May 15 11:53:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W050919S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 10 03:41:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	817649	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	1465548	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	636080	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	136360	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	759487	87.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	174.74%	
35) Dibromofluoromethane	7.88	113	639058	74.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	149.04%	
50) Toluene-d8	10.32	98	955553	27.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	54.72%	
62) 4-Bromofluorobenzene	12.62	95	219550	17.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	35.72%	
Target Compounds						
8) Diethyl Ether	3.68	74	14600	3.189	ug/l	90
16) Acetone	4.12	43	183504	80.977	ug/l	99
20) Methylene Chloride	4.91	84	166232	17.587	ug/l	91
25) 2-Butanone	7.18	43	28283	8.158	ug/l	98
52) Toluene	10.39	92	88238	3.317	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW051119\
Data File : VW010391.D
Acq On : 11 May 2019 19:57
Operator : SY/VA
Sample : K2719-04
Misc : 4.08G/5ML/MSVOA W/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
P026-SS008-0206-02

Quant Time: May 15 11:53:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W050919S.M
Quant Title : SW846 8260
QLast Update : Fri May 10 03:41:21 2019
Response via : Initial Calibration

