


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW051520\
 Data File : VW015427.D
 Acq On : 15 May 2020 12:00
 Operator : SY/VA
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W050620S.M
 Title : SW846 8260
 Last Update : Tue May 05 18:41:17 2020



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.0	23109	PASS
75	95	30	60	53.1	58560	PASS
95	95	100	100	100.0	110221	PASS
96	95	5	9	6.7	7341	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	77.3	85221	PASS
175	174	5	9	7.5	6423	PASS
176	174	95	101	96.9	82570	PASS
177	176	5	9	7.0	5788	PASS