

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051622\
 Data File : VW023046.D
 Acq On : 16 May 2022 12:20
 Operator : SY/VA
 Sample : N2821-17RE
 Misc : 4.80g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EXLD7

Quant Time: May 17 01:26:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue May 17 01:24:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	346072	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	319281	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	137285	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	109380	15.405	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.640%
7) Chloroethane-d5	2.885	69	92377	19.998	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.000%
11) 1,1-Dichloroethene-d2	4.019	63	113501	12.372	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	49.480%
21) 2-Butanone-d5	7.080	46	75353	55.818	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	111.640%
24) Chloroform-d	7.647	84	224143	22.107	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	88.440%
26) 1,2-Dichloroethane-d4	8.305	65	134530	24.192	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	96.760%
32) Benzene-d6	8.274	84	412593	22.420	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.680%
36) 1,2-Dichloropropane-d6	9.274	67	129761	25.101	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.400%
41) Toluene-d8	10.323	98	367104	20.650	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	82.600%
43) trans-1,3-Dichloroprop...	10.579	79	52141	21.906	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.640%
47) 2-Hexanone-d5	10.920	63	62754	63.590	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	127.180%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	133774	27.764	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	111.040%
66) 1,2-Dichlorobenzene-d4	13.853	152	114268	23.811	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.240%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051622\
Data File : VW023046.D
Acq On : 16 May 2022 12:20
Operator : SY/VA
Sample : N2821-17RE
Misc : 4.80g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EXLD7

Quant Time: May 17 01:26:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Tue May 17 01:24:30 2022
Response via : Initial Calibration

