

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051622\
 Data File : VW023062.D
 Acq On : 16 May 2022 19:52
 Operator : SY/VA
 Sample : VW0516SBL02
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK506

Quant Time: May 17 01:31:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue May 17 01:24:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	297115	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	296295	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	152015	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	160654	26.355	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 105.440%	
7) Chloroethane-d5	2.885	69	112665	28.409	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 113.640%	
11) 1,1-Dichloroethene-d2	4.019	63	148077	18.800	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 75.200%	
21) 2-Butanone-d5	7.074	46	73362	63.298	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 126.600%	
24) Chloroform-d	7.647	84	242809	27.894	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 111.560%	
26) 1,2-Dichloroethane-d4	8.305	65	142776	29.906	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 119.640%	
32) Benzene-d6	8.275	84	484177	28.351	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 113.400%	
36) 1,2-Dichloropropane-d6	9.275	67	139807	29.143	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 116.560%	
41) Toluene-d8	10.323	98	457928	27.757	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 111.040%	
43) trans-1,3-Dichloroprop...	10.573	79	62731	28.399	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 113.600%	
47) 2-Hexanone-d5	10.921	63	60893	66.491	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 132.980%	
56) 1,1,2,2-Tetrachloroeth...	12.689	84	133542	29.866	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 119.480%	
66) 1,2-Dichlorobenzene-d4	13.853	152	158670	29.859	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 119.440%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051622\
Data File : VW023062.D
Acq On : 16 May 2022 19:52
Operator : SY/VA
Sample : VW0516SBL02
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK506

Quant Time: May 17 01:31:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Tue May 17 01:24:30 2022
Response via : Initial Calibration

