

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051622\
 Data File : VW023066.D
 Acq On : 16 May 2022 21:29
 Operator : SY/VA
 Sample : N2843-07
 Misc : 6.46g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW4H7

Quant Time: May 17 03:12:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue May 17 02:41:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	229266	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	185659	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	60610	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	82133	17.461	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	69.840%
7) Chloroethane-d5	2.885	69	65504	21.405	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.640%
11) 1,1-Dichloroethene-d2	4.019	63	77298	12.718	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	50.880%
21) 2-Butanone-d5	7.086	46	42534	47.560	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.120%
24) Chloroform-d	7.653	84	156603	23.315	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.240%
26) 1,2-Dichloroethane-d4	8.305	65	94086	25.539	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	102.160%
32) Benzene-d6	8.275	84	271401	25.362	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	101.440%
36) 1,2-Dichloropropane-d6	9.274	67	88599	29.474	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	117.880%
41) Toluene-d8	10.323	98	224427	21.710	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.840%
43) trans-1,3-Dichloroprop...	10.579	79	28647	20.697	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	82.800%
47) 2-Hexanone-d5	10.927	63	35648	62.122	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	124.240%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	91137	32.528	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	130.120%#
66) 1,2-Dichlorobenzene-d4	13.853	152	54459	25.704	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	102.800%
Target Compounds					Qvalue	
48) 2-Hexanone	10.945	43	4325	3.880	ug/L #	41


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051622\
Data File : VW023066.D
Acq On : 16 May 2022 21:29
Operator : SY/VA
Sample : N2843-07
Misc : 6.46g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW4H7


Quant Time: May 17 03:12:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Tue May 17 02:41:35 2022
Response via : Initial Calibration





#48
 2-Hexanone
 Concen: 3.880 ug/L
 RT: 10.945 min Scan# 14
 Delta R.T. -0.024 min
 Lab File: VW023066.D
 Acq: 16 May 2022 21:29

Instrument :
 MSVOA_W
 ClientSampleId :
 EW4H7



Tgt Ion: 43	Resp: 4325
Ion Ratio	Lower Upper
43	100
58	0.0 45.9 68.9#
57	8.9 16.2 24.2#
100	0.0 10.7 16.1#

