

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051622\
 Data File : VW023055.D
 Acq On : 16 May 2022 15:59
 Operator : SY/VA
 Sample : N2821-10RE
 Misc : 4.68g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EXLD0RE

Quant Time: May 17 01:28:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue May 17 01:24:30 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		8.842	114	363815	25.000	ug/L	0.00
28) Chlorobenzene-d5		11.628	117	335892	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		13.554	152	133368	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		2.349	65	231882	31.066	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	124.280%	
7) Chloroethane-d5		2.879	69	201492	41.493	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	165.960%#	
11) 1,1-Dichloroethene-d2		4.013	63	228767	23.720	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	94.880%	
21) 2-Butanone-d5		7.074	46	133873	94.331	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	188.660%#	
24) Chloroform-d		7.647	84	449093	42.133	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	168.520%#	
26) 1,2-Dichloroethane-d4		8.305	65	259873	44.453	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	177.800%#	
32) Benzene-d6		8.275	84	825311	42.628	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	170.520%#	
36) 1,2-Dichloropropane-d6		9.274	67	256083	47.088	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	188.360%#	
41) Toluene-d8		10.323	98	746452	39.912	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	159.640%#	
43) trans-1,3-Dichloroprop...		10.579	79	102078	40.765	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	163.040%#	
47) 2-Hexanone-d5		10.921	63	111047	106.962	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	213.920%#	
56) 1,1,2,2-Tetrachloroeth...		12.688	84	255185	50.342	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	201.360%#	
66) 1,2-Dichlorobenzene-d4		13.853	152	214653	46.042	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	184.160%#	
Target Compounds							
46) Tetrachloroethene		10.860	164	6671	1.631	ug/L	Qvalue # 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051622\
Data File : VW023055.D
Acq On : 16 May 2022 15:59
Operator : SY/VA
Sample : N2821-10RE
Misc : 4.68g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EXLD0RE

Quant Time: May 17 01:28:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Tue May 17 01:24:30 2022
Response via : Initial Calibration

