


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW051719\
 Data File : VW010416.D
 Acq On : 17 May 2019 16:05
 Operator : SY/VA
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W051719S.M
 Title : SW846 8260
 Last Update : Sat May 18 01:59:40 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.5	82794	PASS
75	95	30	60	47.3	238017	PASS
95	95	100	100	100.0	502790	PASS
96	95	5	9	6.6	33189	PASS
173	174	0.00	2	1.3	4497	PASS
174	95	50	100	69.9	351530	PASS
175	174	5	9	7.6	26754	PASS
176	174	95	101	97.8	343722	PASS
177	176	5	9	6.8	23234	PASS