

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051921\
 Data File : VW019023.D
 Acq On : 18 May 2021 16:42
 Operator : SY/VA
 Sample : VIBLK
 Misc : 5.00G/5ML/MSVOA_W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK

Quant Time: May 19 07:14:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W051921S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 07:04:56 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.951	168	400162	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.842	114	671499	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.634	117	591919	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.560	152	292307	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.305	65	230453	56.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.92%	
35) Dibromofluoromethane	7.884	113	206025	55.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.20%	
50) Toluene-d8	10.323	98	834608	56.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.44%	
62) 4-Bromofluorobenzene	12.615	95	288396	54.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.72%	
Target Compounds						
16) Acetone	4.141	43	10134	15.79	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051921\
Data File : VW019023.D
Acq On : 18 May 2021 16:42
Operator : SY/VA
Sample : VIBLK
Misc : 5.00G/5ML/MSVOA_W/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK

Quant Time: May 19 07:14:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W051921S.M
Quant Title : SW846 8260
QLast Update : Wed May 19 07:04:56 2021
Response via : Initial Calibration

