


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW052019\
Data File : VW010446.D
Acq On : 20 May 2019 12:55
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W052019S.M
Title : SW846 8260
Last Update : Tue May 21 06:38:43 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1743

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.3	84440	PASS
75	95	30	60	47.6	246466	PASS
95	95	100	100	100.0	518052	PASS
96	95	5	9	6.7	34520	PASS
173	174	0.00	2	1.3	4711	PASS
174	95	50	100	70.1	363050	PASS
175	174	5	9	7.6	27458	PASS
176	174	95	101	97.8	355093	PASS
177	176	5	9	6.3	22354	PASS