

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW052119\
 Data File : VW010490.D
 Acq On : 21 May 2019 16:26
 Operator : SY/VA
 Sample : K2935-04RE
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 PURGE-WATER-MUDRE

Quant Time: May 28 05:48:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W052019S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 21 06:38:43 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	71683	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	136842	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	98636	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	24927	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	32875	44.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.86%	
35) Dibromofluoromethane	7.88	113	32091	38.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.30%	
50) Toluene-d8	10.32	98	100110	31.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	63.12%	
62) 4-Bromofluorobenzene	12.62	95	28842	22.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	45.66%	
Target Compounds						
16) Acetone	4.14	43	4342	28.750	ug/l	92
17) Carbon Disulfide	4.38	76	3918	1.735	ug/l	95
30) Chloroform	7.68	83	19611	13.557	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW052119\
Data File : VW010490.D
Acq On : 21 May 2019 16:26
Operator : SY/VA
Sample : K2935-04RE
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
PURGE-WATER-MUDRE

Quant Time: May 28 05:48:20 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W052019S.M
Quant Title : SW846 8260
QLast Update : Tue May 21 06:38:43 2019
Response via : Initial Calibration

