


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW052119\
Data File : VW010476.D
Acq On : 21 May 2019 09:44
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W052019S.M
Title : SW846 8260
Last Update : Tue May 21 06:38:43 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.2	14604	PASS
75	95	30	60	47.5	42784	PASS
95	95	100	100	100.0	90149	PASS
96	95	5	9	6.6	5991	PASS
173	174	0.00	2	1.3	839	PASS
174	95	50	100	69.3	62493	PASS
175	174	5	9	7.7	4823	PASS
176	174	95	101	95.3	59544	PASS
177	176	5	9	6.6	3933	PASS