

Data Path : W:\HPCHEM1\MSVOA_W\DATA\VW052218\
 Data File : VW002789.D
 Acq On : 22 May 2018 17:12
 Operator : JC/SY
 Sample : J2959-04
 Misc : 5.36G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 23 05:40:47 2018
 Quant Method : W:\HPCHEM1\MSVOA_W\METHOD\SOM2WLM052218S.M
 Quant Title : VOC Analysis
 QLast Update : Tue May 22 12:32:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	251805	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.64	117	221039	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	93474	25.00	ug/L	0.00

System Monitoring Compounds


4) Vinyl Chloride-d3	2.34	65	77455	23.98	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	95.92%
7) Chloroethane-d5	2.89	69	66842	23.14	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.56%
10) 1,1-Dichloroethene-d2	4.00	63	95985	16.97	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.88%
20) 2-Butanone-d5	7.08	46	69449	52.72	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	105.44%
24) Chloroform-d	7.65	84	140088	23.21	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.84%
26) 1,2-Dichloroethane-d4	8.31	65	88931	24.40	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.60%
29) Benzene-d6	8.28	84	263161	25.03	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	100.12%
33) 1,2-Dichloropropane-d6	9.28	67	85442	25.15	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.60%
37) Toluene-d8	10.33	98	240287	23.16	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.64%
38) trans-1,3-Dichloropropene-	10.59	79	38036	23.41	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	93.64%
39) 2-Hexanone-d5	10.93	63	53457	55.46	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.92%
48) 1,1,2,2-Tetrachloroethane-	12.70	84	95594	25.67	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	102.68%
61) 1,2-Dichlorobenzene-d4	13.87	152	80980	23.57	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	94.28%


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
13) Acetone	4.13	43	2570	2.21	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA_W\DATA\VW052218\
Data File : VW002789.D
Acq On : 22 May 2018 17:12
Operator : JC/SY
Sample : J2959-04
Misc : 5.36G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 23 05:40:47 2018
Quant Method : W:\HPCHEM1\MSVOA_W\METHOD\SOM2WLM052218S.M
Quant Title : VOC Analysis
QLast Update : Tue May 22 12:32:39 2018
Response via : Initial Calibration





#13

Acetone

Concen: 2.21 ug/L

RT: 4.13 min Scan# 417

Delta R.T. 0.01 min

Lab File: VW002789.D

Acq: 22 May 2018 17:12

Tgt Ion: 43 Resp: 2570

Ion Ratio Lower Upper

43 100

58 26.3 0.0 59.8

