


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW052219\
Data File : VW010499.D
Acq On : 22 May 2019 09:47
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W052019S.M
Title : SW846 8260
Last Update : Tue May 21 06:38:43 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.7	15239	PASS
75	95	30	60	47.1	45709	PASS
95	95	100	100	100.0	97114	PASS
96	95	5	9	6.8	6589	PASS
173	174	0.00	2	1.0	686	PASS
174	95	50	100	69.4	67381	PASS
175	174	5	9	7.8	5251	PASS
176	174	95	101	95.0	63994	PASS
177	176	5	9	6.7	4284	PASS