

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW052220\
 Data File : VW015454.D
 Acq On : 22 May 2020 16:15
 Operator : SY/VA
 Sample : VIBLK10
 Misc : 5.00G/10.0mL/MSVOA W/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK10

Quant Time: May 23 05:36:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM050420S.M
 Quant Title : VOC Analysis
 QLast Update : Sat May 23 03:53:37 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	353131	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	320436	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	152747	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	74351	16.83	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	67.32%
7) Chloroethane-d5	2.89	69	83289	19.56	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.24%
10) 1,1-Dichloroethene-d2	4.01	63	131923	14.68	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	58.72%
20) 2-Butanone-d5	7.08	46	54838	46.47	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.94%
24) Chloroform-d	7.65	84	193101	21.65	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.60%
26) 1,2-Dichloroethane-d4	8.30	65	114495	23.16	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	92.64%
29) Benzene-d6	8.27	84	391151	22.31	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.24%
33) 1,2-Dichloropropane-d6	9.27	67	118284	22.26	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	89.04%
37) Toluene-d8	10.32	98	372473	22.09	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.36%
38) trans-1,3-Dichloropropene-	10.58	79	50916	21.05	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.20%
39) 2-Hexanone-d5	10.93	63	44364	45.20	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.40%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	90935	22.72	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	90.88%
61) 1,2-Dichlorobenzene-d4	13.85	152	125249	23.39	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.56%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW052220\
Data File : VW015454.D
Acq On : 22 May 2020 16:15
Operator : SY/VA
Sample : VIBLK10
Misc : 5.00G/10.0mL/MSVOA W/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK10

Quant Time: May 23 05:36:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM050420S.M
Quant Title : VOC Analysis
QLast Update : Sat May 23 03:53:37 2020
Response via : Initial Calibration

