

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW052319\
 Data File : VW010540.D
 Acq On : 23 May 2019 19:55
 Operator : SY/VA
 Sample : K2976-12
 Misc : 6.32G/5ML/MSVOA W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EP1-SB-MW-ER(17-20)

Quant Time: May 24 08:44:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W052019S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 21 06:38:43 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	123653	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	228510	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	201538	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	111336	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	70979	55.61	ug/l	0.00
Spiked Amount			Recovery	=	111.22%	
35) Dibromofluoromethane	7.88	113	67798	48.90	ug/l	0.00
Spiked Amount			Recovery	=	97.80%	
50) Toluene-d8	10.32	98	283279	53.48	ug/l	0.00
Spiked Amount			Recovery	=	106.96%	
62) 4-Bromofluorobenzene	12.62	95	187935	89.07	ug/l	0.00
Spiked Amount			Recovery	=	178.14%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.12	43	6815	26.159	ug/l #	86
17) Carbon Disulfide	4.37	76	4649	1.193	ug/l	99
20) Methylene Chloride	4.91	84	4932	Below Cal		95
25) 2-Butanone	7.18	43	424	1.137	ug/l #	64
31) Cyclohexane	7.95	56	3198	1.320	ug/l #	33
36) 1,1-Dichloropropene	7.95	75	9893	4.647	ug/l #	48
38) Carbon Tetrachloride	7.95	117	13153	5.728	ug/l #	16
55) 1,1,2-Trichloroethane	10.76	97	37371	31.624	ug/l #	22
56) Ethyl methacrylate	10.66	69	16334	10.022	ug/l #	14
59) 2-Hexanone	10.94	43	109382	167.330	ug/l #	27
67) Ethyl Benzene	11.73	91	16477	2.173	ug/l	97
73) Isopropylbenzene	12.47	105	77901	8.817	ug/l	100
74) N-amyl acetate	12.26	43	870932	430.184	ug/l #	61
75) 1,1,2,2-Tetrachloroethane	12.69	83	91767	57.106	ug/l #	75
76) 1,2,3-Trichloropropane	12.62	75	42390	37.509	ug/l #	1
78) n-propylbenzene	12.80	91	95681	9.293	ug/l	92
79) 2-Chlorotoluene	12.70	91	15706	2.630	ug/l	85
80) 1,3,5-Trimethylbenzene	12.95	105	741011	99.031	ug/l	99
81) trans-1,4-Dichloro-2-buten	12.62	75	42390	78.282	ug/l #	50
82) 4-Chlorotoluene	12.94	91	101912	16.444	ug/l #	67
83) tert-Butylbenzene	13.21	119	80924	12.381	ug/l	89
84) 1,2,4-Trimethylbenzene	13.26	105	4057570	544.460	ug/l	99
85) sec-Butylbenzene	13.39	105	310140	33.907	ug/l #	76
86) p-Isopropyltoluene	13.51	119	388669	46.946	ug/l	99
89) n-Butylbenzene	13.80	91	79872	10.143	ug/l #	1
90) Hexachloroethane	14.10	117	86354	53.887	ug/l #	15
92) 1,2-Dibromo-3-Chloropropan	14.50	75	1005	3.363	ug/l #	12
95) Naphthalene	15.37	128	138791	27.495	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW052319\
Data File : VW010540.D
Acq On : 23 May 2019 19:55
Operator : SY/VA
Sample : K2976-12
Misc : 6.32G/5ML/MSVOA W/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EP1-SB-MW-ER(17-20)

Quant Time: May 24 08:44:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W052019S.M
Quant Title : SW846 8260
QLast Update : Tue May 21 06:38:43 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.95 min Scan# 991

Delta R.T. -0.01 min

Lab File: VW010540.D

Acq: 23 May 2019 19:55

Instrument :

MSVOA_W

ClientSampleId :


EP1-SB-MW-ER(17-20)

Tot Ion:168 Resp: 123653

Ion Ratio Lower Upper

168 100

99 63.9 46.1 69.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VW

7.95

60000

50000

40000

30000



20000

10000

0

Time-->

7.90 8.00



#16

Acetone

Concen: 26.159 ug/l

RT: 4.12 min Scan# 363

Delta R.T. -0.01 min

Lab File: VW010540.D


Acq: 23 May 2019 19:55

Tgt Ion: 43 Resp: 6815

Ion Ratio Lower Upper

43 100

58 25.9 27.0 40.4#



Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 58.00 (57.70 to 58.70): VW

4.12

2500

2000

1500


1000


500

0

Time-->

4.00 4.10 4.20





#17

Carbon Disulfide

Concen: 1.193 ug/l

RT: 4.37 min Scan# 405

Delta R.T. -0.00 min

Lab File: VW010540.D

Acq: 23 May 2019 19:55

Instrument :

MSVOA_W

ClientSampleId :


EP1-SB-MW-ER(17-20)

Tot Ion: 76 Resp: 4649




Ion Ratio Lower Upper

76 100

78 9.3 7.1 10.7



Abundance Ion 75.90 (75.60 to 76.60): VW
Ion 77.90 (77.60 to 78.60): VW



#20

Methylene Chloride

Concen: Below Cal

RT: 4.91 min Scan# 493

Delta R.T. -0.01 min

Lab File: VW010540.D

Acq: 23 May 2019 19:55

Tgt Ion: 84 Resp: 4932


Ion Ratio Lower Upper

84 100

49 102.1 87.2 130.8

51 33.2 27.7 41.5

86 62.5 51.4 77.0






Abundance Ion 83.90 (83.60 to 84.60): VW

Ion 48.90 (48.60 to 49.60): VW

Ion 50.90 (50.60 to 51.60): VW

Ion 85.90 (85.60 to 86.60): VW





#25

2-Butanone

Concen: 1.137 ug/l

RT: 7.18 min Scan# 866

Delta R.T. 0.01 min

Lab File: VW010540.D

Acq: 23 May 2019 19:55

Instrument :

MSVOA_W

ClientSampleId :


EP1-SB-MW-ER(17-20)

Tot Ion: 43 Resp: 424

Ion Ratio Lower Upper

43 100

72 49.5 24.1 36.1#



Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 72.00 (71.70 to 72.70): VW

7.18

250

200

150



100

50

0

Time-->

7.15 7.20



#31

Cyclohexane

Concen: 1.320 ug/l

RT: 7.95 min Scan# 992

Delta R.T. -0.00 min

Lab File: VW010540.D

Acq: 23 May 2019 19:55


Tgt Ion: 56 Resp: 3198

Ion Ratio Lower Upper

56 100

69 144.3 28.2 42.2#

84 116.0 75.8 113.8#



Abundance Ion 56.00 (55.70 to 56.70): VW

Ion 69.00 (68.70 to 69.70): VW

Ion 84.00 (83.70 to 84.70): VW

3000

2500

2000

1500


1000


500

0

Time-->

7.85 7.90 7.95 8.00









#33
 1,2-Dichloroethane-d4
 Concen: 55.613 ug/l
 RT: 8.31 min Scan# 1050
 Delta R.T. -0.00 min
 Lab File: VW010540.D
 Acq: 23 May 2019 19:55




Instrument :
 MSVOA_W
 ClientSampleId :
 EP1-SB-MW-ER(17-20)


Tot Ion: 65 Resp: 70979
 Ion Ratio Lower Upper
 65 100
 67 51.4 0.0 103.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.84 min Scan# 1138
 Delta R.T. -0.00 min
 Lab File: VW010540.D
 Acq: 23 May 2019 19:55

Tgt Ion: 114 Resp: 228510
 Ion Ratio Lower Upper
 114 100
 63 18.5 0.0 36.0
 88 16.3 0.0 32.0









#35
 Dibromofluoromethane
 Concen: 48.897 ug/l
 RT: 7.88 min Scan# 981
 Delta R.T. -0.00 min
 Lab File: VW010540.D
 Acq: 23 May 2019 19:55




Instrument :
 MSVOA_W
 ClientSampleId :
 EP1-SB-MW-ER(17-20)


Tot Ion:113 Resp: 67798
 Ion Ratio Lower Upper
 113 100
 111 104.0 82.2 123.4
 192 17.6 14.3 21.5



#36
 1,1-Dichloropropene
 Concen: 4.647 ug/l
 RT: 7.95 min Scan# 991
 Delta R.T. -0.13 min
 Lab File: VW010540.D
 Acq: 23 May 2019 19:55

Tgt Ion: 75 Resp: 9893
 Ion Ratio Lower Upper
 75 100
 110 9.3 19.2 57.6#
 77 0.0 25.2 37.8#









#38
Carbon Tetrachloride
Concen: 5.728 ug/l
RT: 7.95 min Scan# 992
Delta R.T. -0.12 min
Lab File: VW010540.D
Acq: 23 May 2019 19:55

Instrument :
MSVOA_W
ClientSampleId :
EP1-SB-MW-ER(17-20)

Tot Ion:117 Resp: 13153
Ion Ratio Lower Upper
117 100
119 5.7 77.4 116.0#
121 0.0 25.0 37.6#




Abundance Ion 116.80 (116.50 to 117.50): V
Ion 118.80 (118.50 to 119.50): V
Ion 120.80 (120.50 to 121.50): V






#50
Toluene-d8
Concen: 53.480 ug/l
RT: 10.32 min Scan# 1381
Delta R.T. -0.00 min
Lab File: VW010540.D
Acq: 23 May 2019 19:55

Tgt Ion: 98 Resp: 283279
Ion Ratio Lower Upper
98 100
100 66.3 53.7 80.5



Abundance Ion 98.00 (97.70 to 98.70): VW
Ion 100.00 (99.70 to 100.70): VW









#55
 1,1,2-Trichloroethane
 Concen: 31.624 ug/l
 RT: 10.76 min Scan# 1453
 Delta R.T. -0.02 min
 Lab File: VW010540.D
 Acq: 23 May 2019 19:55




Instrument :
 MSVOA_W
 ClientSampleId :
 EP1-SB-MW-ER(17-20)


Tot Ion: 97 Resp: 37371
 Ion Ratio Lower Upper
 97 100
 83 8.0 66.1 99.1#
 85 8.1 44.2 66.4#
 99 0.3 51.1 76.7#



#56
 Ethyl methacrylate
 Concen: 10.022 ug/l
 RT: 10.66 min Scan# 1437
 Delta R.T. 0.02 min
 Lab File: VW010540.D
 Acq: 23 May 2019 19:55

Tgt Ion: 69 Resp: 16334
 Ion Ratio Lower Upper
 69 100
 41 133.5 47.7 71.5#
 39 67.5 25.3 37.9#









#59
2-Hexanone
Concen: 167.330 ug/l
RT: 10.94 min Scan# 1482
Delta R.T. -0.03 min
Lab File: VW010540.D
Acq: 23 May 2019 19:55




Instrument :
MSVOA_W
ClientSampleId :
EP1-SB-MW-ER(17-20)


Tot Ion: 43 Resp: 109382
Ion Ratio Lower Upper
43 100
58 3.6 28.9 86.8#



#62
4-Bromofluorobenzene
Concen: 89.075 ug/l
RT: 12.62 min Scan# 1757
Delta R.T. -0.01 min
Lab File: VW010540.D
Acq: 23 May 2019 19:55

Tgt Ion: 95 Resp: 187935
Ion Ratio Lower Upper
95 100
174 37.6 0.0 148.6
176 36.2 0.0 144.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.63 min Scan# 1596
Delta R.T. -0.00 min
Lab File: VW010540.D
Acq: 23 May 2019 19:55




Instrument :
MSVOA_W
ClientSampleId :
EP1-SB-MW-ER(17-20)


Tot Ion	Ratio	Lower	Upper
117	100		
82	48.8	42.8	64.2
119	32.6	26.3	39.5



#67
Ethyl Benzene
Concen: 2.173 ug/l
RT: 11.73 min Scan# 1612
Delta R.T. -0.00 min
Lab File: VW010540.D
Acq: 23 May 2019 19:55

Tgt Ion	Ratio	Lower	Upper
91	100		
106	34.1	25.9	38.9





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.56 min Scan# 1912

Delta R.T. -0.00 min

Lab File: VW010540.D

Acq: 23 May 2019 19:55

Instrument :

MSVOA_W

ClientSampleId :

EP1-SB-MW-ER(17-20)


Tot Ion:152 Resp: 111336

Ion Ratio Lower Upper

152 100

115 0.0 30.3 91.0#




150 117.7 0.0 349.8



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#73

Isopropylbenzene

Concen: 8.817 ug/l

RT: 12.47 min Scan# 1733

Delta R.T. -0.00 min

Lab File: VW010540.D


Acq: 23 May 2019 19:55

Tgt Ion:105 Resp: 77901

Ion Ratio Lower Upper



105 100


120 27.7 13.9 41.7



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#74

N-amvl acetate

Concen: 430.184 ug/l

RT: 12.26 min Scan# 1698

Delta R.T. -0.02 min

Lab File: VW010540.D

Acq: 23 May 2019 19:55

Instrument :

MSVOA_W

ClientSampleId :

EP1-SB-MW-ER(17-20)

Tot Ion: 43 Resp: 870932





Ion Ratio Lower Upper

43 100

70 18.7 37.1 55.7#

55 13.9 19.6 29.4#

61 0.0 20.7 31.1#



#75

1,1,2,2-Tetrachloroethane

Concen: 57.106 ug/l

RT: 12.69 min Scan# 1770

Delta R.T. -0.02 min

Lab File: VW010540.D

Acq: 23 May 2019 19:55




Tgt Ion: 83 Resp: 91767


Ion Ratio Lower Upper

83 100

131 0.0 5.6 16.8#

85 83.5 32.4 97.0





#76

1,2,3-Trichloropropane

Concen: 37.509 ug/l

RT: 12.62 min Scan# 1758

Delta R.T. -0.15 min

Lab File: VW010540.D

Acq: 23 May 2019 19:55

Instrument :

MSVOA_W

ClientSampleId :





EP1-SB-MW-ER(17-20)

Tot Ion: 75 Resp: 42390

Ion Ratio Lower Upper

75 100

77 26.5 170.0 510.1#



#78

n-propylbenzene

Concen: 9.293 ug/l

RT: 12.80 min Scan# 1788

Delta R.T. -0.00 min

Lab File: VW010540.D




Acq: 23 May 2019 19:55


Tgt Ion: 91 Resp: 95681

Ion Ratio Lower Upper

91 100

120 20.5 12.3 36.8









#79
 2-Chlorotoluene
 Concen: 2.630 ug/l
 RT: 12.70 min Scan# 1771
 Delta R.T. -0.20 min
 Lab File: VW010540.D
 Acq: 23 May 2019 19:55




Instrument :
 MSVOA_W
 ClientSampleId :
 EP1-SB-MW-ER(17-20)


Tot Ion: 91 Resp: 15706
 Ion Ratio Lower Upper
 91 100
 126 26.0 17.4 52.3



#80
 1,3,5-Trimethylbenzene
 Concen: 99.031 ug/l
 RT: 12.95 min Scan# 1812
 Delta R.T. 0.01 min
 Lab File: VW010540.D
 Acq: 23 May 2019 19:55

Tgt Ion:105 Resp: 741011
 Ion Ratio Lower Upper
 105 100
 120 49.9 25.2 75.6





#81

trans-1,4-Dichloro-2-butene

Concen: 78.282 ug/l

RT: 12.62 min Scan# 1758

Delta R.T. 0.11 min

Lab File: VW010540.D

Acq: 23 May 2019 19:55

Instrument :

MSVOA_W

ClientSampleId :

EP1-SB-MW-ER(17-20)





Tot Ion: 75 Resp: 42390

Ion Ratio Lower Upper

75 100

53 51.8 69.8 104.8#

89 1.1 39.5 59.3#



#82

4-Chlorotoluene

Concen: 16.444 ug/l

RT: 12.94 min Scan# 1811

Delta R.T. -0.05 min

Lab File: VW010540.D




Acq: 23 May 2019 19:55


Tgt Ion: 91 Resp: 101912

Ion Ratio Lower Upper

91 100

126 15.8 17.5 52.6#









#83
tert-Butylbenzene
Concen: 12.381 ug/l
RT: 13.21 min Scan# 1855
Delta R.T. 0.01 min
Lab File: VW010540.D
Acq: 23 May 2019 19:55




Instrument :
MSVOA_W
ClientSampleId :
EP1-SB-MW-ER(17-20)


Tot Ion:119 Resp: 80924
Ion Ratio Lower Upper
119 100
91 78.5 33.1 99.5
134 26.5 13.9 41.6



#84
1,2,4-Trimethylbenzene
Concen: 544.460 ug/l
RT: 13.26 min Scan# 1862
Delta R.T. -0.00 min
Lab File: VW010540.D
Acq: 23 May 2019 19:55

Tgt Ion:105 Resp: 4057570
Ion Ratio Lower Upper
105 100
120 47.3 23.4 70.2









#85
 sec-Butylbenzene
 Concen: 33.907 ug/l
 RT: 13.39 min Scan# 1884
 Delta R.T. 0.01 min
 Lab File: VW010540.D
 Acq: 23 May 2019 19:55




Instrument :
 MSVOA_W
 ClientSampleId :
 EP1-SB-MW-ER(17-20)


Tot Ion:105 Resp: 310140
 Ion Ratio Lower Upper
 105 100
 134 32.3 10.5 31.6#



#86
 p-Isopropyltoluene
 Concen: 46.946 ug/l
 RT: 13.51 min Scan# 1903
 Delta R.T. 0.01 min
 Lab File: VW010540.D
 Acq: 23 May 2019 19:55

Tgt Ion:119 Resp: 388669
 Ion Ratio Lower Upper
 119 100
 134 27.1 13.6 40.7
 91 24.8 11.9 35.7









#89
n-Butylbenzene
Concen: 10.143 ug/l
RT: 13.80 min Scan# 1951
Delta R.T. -0.03 min
Lab File: VW010540.D
Acq: 23 May 2019 19:55




Instrument :
MSVOA_W
ClientSampleId :
EP1-SB-MW-ER(17-20)


Tot Ion: 91 Resp: 79872
Ion Ratio Lower Upper
91 100
92 0.0 27.7 83.1#
134 378.9 13.9 41.6#



#90
Hexachloroethane
Concen: 53.887 ug/l
RT: 14.10 min Scan# 2001
Delta R.T. 0.01 min
Lab File: VW010540.D
Acq: 23 May 2019 19:55

Tgt Ion: 117 Resp: 86354
Ion Ratio Lower Upper
117 100
201 0.0 34.6 103.8#









#92
 1,2-Dibromo-3-Chloropropane
 Concen: 3.363 ug/l
 RT: 14.50 min Scan# 2066
 Delta R.T. 0.01 min
 Lab File: VW010540.D
 Acq: 23 May 2019 19:55

Instrument :
 MSVOA_W
 ClientSampleId :
 EP1-SB-MW-ER(17-20)

Tot Ion: 75 Resp: 1005
 Ion Ratio Lower Upper
 75 100
 155 25.5 43.0 129.0#
 157 0.0 55.7 167.1#



#95
 Naphthalene
 Concen: 27.495 ug/l
 RT: 15.37 min Scan# 2209
 Delta R.T. -0.00 min
 Lab File: VW010540.D
 Acq: 23 May 2019 19:55

Tgt Ion: 128 Resp: 138791
 Ion Ratio Lower Upper
 128 100
 127 13.1 10.6 15.8
 129 10.5 8.9 13.3

