

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052421\
 Data File : VW019067.D
 Acq On : 24 May 2021 14:24
 Operator : SY/VA
 Sample : VSTDICV050
 Misc : 5.00G/5ML/MSVOA_W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ICVVW052421

Quant Time: May 25 04:07:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W052421S.M
 Quant Title : SW846 8260
 QLast Update : Mon May 24 14:01:57 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	0.352	0.347	1.4	103	0.00
3 P	Chloromethane	0.389	0.380	2.3	110	0.00
4 C	Vinyl Chloride	0.610	0.586	3.9#	100	0.00
5 T	Bromomethane	0.494	0.457	7.5	102	0.00
6 T	Chloroethane	0.405	0.387	4.4	102	0.00
7 T	Trichlorofluoromethane	0.354	0.334	5.6	98	0.00
8 T	Diethyl Ether	0.279	0.275	1.4	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.535	0.511	4.5	100	0.00
10 T	Methyl Iodide	0.759	0.736	3.0	103	0.00
11 T	Tert butyl alcohol	0.031	0.031	0.0	109	0.00
12 CM	1,1-Dichloroethene	0.557	0.532	4.5#	101	0.00
13 T	Acrolein	0.038	0.034	10.5	107	0.00
14 T	Allyl chloride	0.744	0.714	4.0	101	0.00
15 T	Acrylonitrile	0.103	0.106	-2.9	107	0.00
16 T	Acetone	0.096	0.114	-18.8	127	0.00
17 T	Carbon Disulfide	1.505	1.448	3.8	101	0.00
18 T	Methyl Acetate	0.261	0.251	3.8	104	0.00
19 T	Methyl tert-butyl Ether	0.677	0.677	0.0	105	0.00
20 T	Methylene Chloride	0.700	0.658	6.0	117	0.00
21 T	trans-1,2-Dichloroethene	0.621	0.598	3.7	102	0.00
22 T	Diisopropyl ether	1.440	1.383	4.0	100	0.00
23 T	Vinyl Acetate	0.950	0.958	-0.8	104	0.00
24 P	1,1-Dichloroethane	1.116	1.065	4.6	102	0.00
25 T	2-Butanone	0.142	0.151	-6.3	111	0.00
26 T	2,2-Dichloropropane	0.638	0.614	3.8	104	0.00
27 T	cis-1,2-Dichloroethene	0.682	0.656	3.8	101	0.00
28 T	Bromochloromethane	0.442	0.426	3.6	105	0.00
29 T	Tetrahydrofuran	0.081	0.084	-3.7	108	0.00
30 C	Chloroform	1.125	1.074	4.5#	102	0.00
31 T	Cyclohexane	0.949	0.847	10.7	100	0.00
32 T	1,1,1-Trichloroethane	0.865	0.836	3.4	102	0.00
33 S	1,2-Dichloroethane-d4	0.627	0.597	4.8	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	107	0.00
35 S	Dibromofluoromethane	0.340	0.322	5.3	95	0.00
36 T	1,1-Dichloropropene	0.558	0.541	3.0	102	0.00
37 T	Ethyl Acetate	0.202	0.210	-4.0	107	0.00
38 T	Carbon Tetrachloride	0.495	0.490	1.0	103	0.00
39 T	Methylcyclohexane	0.620	0.601	3.1	100	0.00
40 TM	Benzene	1.521	1.472	3.2	102	0.00
41 T	Methacrylonitrile	0.118	0.121	-2.5	105	0.00
42 TM	1,2-Dichloroethane	0.449	0.442	1.6	103	0.00
43 T	Isopropyl Acetate	0.389	0.406	-4.4	105	0.00
44 TM	Trichloroethene	0.400	0.387	3.3	101	0.00
45 C	1,2-Dichloropropane	0.369	0.357	3.3#	101	0.00
46 T	Dibromomethane	0.192	0.195	-1.6	104	0.00
47 T	Bromodichloromethane	0.495	0.493	0.4	102	0.00
48 T	Methyl methacrylate	0.182	0.187	-2.7	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052421\
 Data File : VW019067.D
 Acq On : 24 May 2021 14:24
 Operator : SY/VA
 Sample : VSTDICV050
 Misc : 5.00G/5ML/MSVOA_W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ICVVW052421

Quant Time: May 25 04:07:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W052421S.M
 Quant Title : SW846 8260
 QLast Update : Mon May 24 14:01:57 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	105	0.00
50 S	Toluene-d8	1.386	1.346	2.9	98	0.00
51 T	4-Methyl-2-Pentanone	0.198	0.208	-5.1	107	0.00
52 CM	Toluene	0.969	0.955	1.4#	103	0.00
53 T	t-1,3-Dichloropropene	0.493	0.505	-2.4	103	0.00
54 T	cis-1,3-Dichloropropene	0.584	0.593	-1.5	102	0.00
55 T	1,1,2-Trichloroethane	0.272	0.270	0.7	103	0.00
56 T	Ethyl methacrylate	0.341	0.359	-5.3	105	0.00
57 T	1,3-Dichloropropane	0.485	0.486	-0.2	103	0.00
58 T	2-Chloroethyl Vinyl ether	0.172	0.176	-2.3	104	0.00
59 T	2-Hexanone	0.133	0.148	-11.3	111	0.00
60 T	Dibromochloromethane	0.304	0.308	-1.3	104	0.00
61 T	1,2-Dibromoethane	0.256	0.264	-3.1	105	0.00
62 S	4-Bromofluorobenzene	0.489	0.484	1.0	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64 T	Tetrachloroethene	0.359	0.338	5.8	99	0.00
65 PM	Chlorobenzene	1.113	1.074	3.5	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.381	-0.3	101	0.00
67 C	Ethyl Benzene	2.109	2.047	2.9#	100	0.00
68 T	m/p-Xylenes	0.774	0.776	-0.3	103	0.00
69 T	o-Xylene	0.731	0.726	0.7	102	0.00
70 T	Styrene	1.221	1.238	-1.4	101	0.00
71 P	Bromoform	0.183	0.192	-4.9	107	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	4.081	4.001	2.0	104	0.00
74 T	N-amyl acetate	0.820	0.854	-4.1	106	0.00
75 P	1,1,2,2-Tetrachloroethane	0.696	0.703	-1.0	105	0.00
76 T	1,2,3-Trichloropropane	0.477	0.464	2.7	104	0.00
77 T	Bromobenzene	0.872	0.862	1.1	105	0.00
78 T	n-propylbenzene	5.056	4.820	4.7	100	0.00
79 T	2-Chlorotoluene	2.850	2.756	3.3	102	0.00
80 T	1,3,5-Trimethylbenzene	3.440	3.391	1.4	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.211	0.222	-5.2	106	0.00
82 T	4-Chlorotoluene	3.004	2.901	3.4	101	0.00
83 T	tert-Butylbenzene	2.932	2.876	1.9	101	0.00
84 T	1,2,4-Trimethylbenzene	3.434	3.374	1.7	100	0.00
85 T	sec-Butylbenzene	4.401	4.308	2.1	102	0.00
86 T	p-Isopropyltoluene	3.749	3.744	0.1	100	0.00
87 T	1,3-Dichlorobenzene	1.805	1.766	2.2	103	0.00
88 T	1,4-Dichlorobenzene	1.778	1.705	4.1	100	0.00
89 T	n-Butylbenzene	3.750	3.687	1.7	100	0.00
90 T	Hexachloroethane	0.592	0.588	0.7	100	0.00
91 T	1,2-Dichlorobenzene	1.572	1.505	4.3	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.111	0.116	-4.5	106	0.00
93 T	1,2,4-Trichlorobenzene	1.086	1.092	-0.6	105	0.00
94 T	Hexachlorobutadiene	0.613	0.605	1.3	100	0.00
95 T	Naphthalene	1.904	1.957	-2.8	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052421\
Data File : VW019067.D
Acq On : 24 May 2021 14:24
Operator : SY/VA
Sample : VSTDICV050
Misc : 5.00G/5ML/MSVOA_W/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
ICVW052421

Quant Time: May 25 04:07:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W052421S.M
Quant Title : SW846 8260
QLast Update : Mon May 24 14:01:57 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.920	0.886	3.7	99	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6