

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052424\
 Data File : VW028733.D
 Acq On : 24 May 2024 14:58
 Operator : SY/MD
 Sample : P2571-01
 Misc : 5.62g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH655

Quant Time: May 25 00:03:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050224SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat May 25 00:01:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	500208	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	395248	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	120403	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	132317	13.930	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	55.720%
7) Chloroethane-d5	2.899	69	123617	16.997	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	68.000%
11) 1,1-Dichloroethene-d2	4.027	65	60445	15.231	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.920%
21) 2-Butanone-d5	7.081	46	47981	19.362	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	38.720%
24) Chloroform-d	7.648	84	316941	20.278	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	81.120%
26) 1,2-Dichloroethane-d4	8.307	65	168980	18.884	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	75.520%
32) Benzene-d6	8.276	84	594657	22.550	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	90.200%
36) 1,2-Dichloropropane-d6	9.276	67	195696	24.662	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	98.640%
41) Toluene-d8	10.325	98	492434	21.086	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	84.360%
43) trans-1,3-Dichloroprop...	10.575	79	48765	15.276	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	61.120%
47) 2-Hexanone-d5	10.922	63	26711	19.517	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	39.040%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	135708	20.395	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	81.600%
66) 1,2-Dichlorobenzene-d4	13.855	152	113028	24.630	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.520%
Target Compounds						Qvalue
16) Methylene chloride	4.923	84	50266	6.498	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052424\
Data File : VW028733.D
Acq On : 24 May 2024 14:58
Operator : SY/MD
Sample : P2571-01
Misc : 5.62g/10mL/MSVOA_W/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH655

Quant Time: May 25 00:03:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050224SMA.M
Quant Title : SFAM01.0
QLast Update : Sat May 25 00:01:38 2024
Response via : Initial Calibration

