

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052424\
 Data File : VW028809.D
 Acq On : 26 May 2024 00:04
 Operator : SY/MD
 Sample : P2590-11
 Misc : 6.27g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BHBL7

Quant Time: May 27 02:26:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050224SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon May 27 02:22:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	174117	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	89773	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	14401	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	58835	17.794	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.160%
7) Chloroethane-d5	2.899	69	61190	24.170	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.680%
11) 1,1-Dichloroethene-d2	4.015	65	23686	17.146	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	68.600%
21) 2-Butanone-d5	7.075	46	68815	79.775	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	159.560%#
24) Chloroform-d	7.648	84	175459	32.250	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	129.000%
26) 1,2-Dichloroethane-d4	8.307	65	82507	26.489	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	105.960%
32) Benzene-d6	8.276	84	294662	49.196	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	196.800%#
36) 1,2-Dichloropropane-d6	9.270	67	123006	68.248	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	273.000%#
41) Toluene-d8	10.325	98	175251	33.040	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	132.160%#
43) trans-1,3-Dichloroprop...	10.575	79	11551	15.931	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	63.720%
47) 2-Hexanone-d5	10.922	63	36010	115.842	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	231.680%#
56) 1,1,2,2-Tetrachloroeth...	12.690	84	98841	65.400	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	261.600%#
66) 1,2-Dichlorobenzene-d4	13.848	152	23098	42.082	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	168.320%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052424\
Data File : VW028809.D
Acq On : 26 May 2024 00:04
Operator : SY/MD
Sample : P2590-11
Misc : 6.27g/10mL/MSVOA_W/SOIL/A
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BHBL7

Quant Time: May 27 02:26:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050224SMA.M
Quant Title : SFAM01.0
QLast Update : Mon May 27 02:22:53 2024
Response via : Initial Calibration

