

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052424\
 Data File : VW028810.D
 Acq On : 26 May 2024 00:28
 Operator : SY/MD
 Sample : P2590-12
 Misc : 4.79g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BHBL8

Quant Time: May 27 02:26:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050224SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon May 27 02:22:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	294307	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	221850	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	61354	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	92221	16.501	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	66.000%
7) Chloroethane-d5	2.899	69	90467	21.141	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.560%
11) 1,1-Dichloroethene-d2	4.027	65	39094	16.743	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	66.960%
21) 2-Butanone-d5	7.075	46	59080	40.520	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	81.040%
24) Chloroform-d	7.648	84	245938	26.744	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	106.960%
26) 1,2-Dichloroethane-d4	8.301	65	138789	26.361	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	105.440%
32) Benzene-d6	8.276	84	418252	28.257	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	113.040%
36) 1,2-Dichloropropane-d6	9.270	67	156082	35.043	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	140.160%#
41) Toluene-d8	10.319	98	300241	22.905	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.640%
43) trans-1,3-Dichloroprop...	10.575	79	32653	18.224	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	72.880%
47) 2-Hexanone-d5	10.922	63	30955	40.296	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.600%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	121705	32.587	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	130.360%#
66) 1,2-Dichlorobenzene-d4	13.848	152	65808	28.142	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	112.560%
Target Compounds						
13) Acetone	4.124	43	12323	9.290	ug/L	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052424\
Data File : VW028810.D
Acq On : 26 May 2024 00:28
Operator : SY/MD
Sample : P2590-12
Misc : 4.79g/10mL/MSVOA_W/SOIL/A
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BHBL8

Quant Time: May 27 02:26:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050224SMA.M
Quant Title : SFAM01.0
QLast Update : Mon May 27 02:22:53 2024
Response via : Initial Calibration

