

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052424\
 Data File : VW028799.D
 Acq On : 25 May 2024 20:04
 Operator : SY/MD
 Sample : P2590-01
 Misc : 4.88g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH6E7

Quant Time: May 27 02:24:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050224SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon May 27 02:22:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	329484	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	205167	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	34081	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	111143	17.764	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.040%
7) Chloroethane-d5	2.899	69	104921	21.901	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.600%
11) 1,1-Dichloroethene-d2	4.021	65	46617	17.833	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	71.320%
21) 2-Butanone-d5	7.081	46	69904	42.825	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.640%
24) Chloroform-d	7.654	84	280476	27.243	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	108.960%
26) 1,2-Dichloroethane-d4	8.307	65	144119	24.451	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.800%
32) Benzene-d6	8.276	84	483870	35.349	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	141.400%#
36) 1,2-Dichloropropane-d6	9.270	67	176124	42.758	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	171.040%#
41) Toluene-d8	10.319	98	337928	27.877	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	111.520%
43) trans-1,3-Dichloroprop...	10.575	79	34540	20.845	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.360%
47) 2-Hexanone-d5	10.922	63	44063	62.023	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	124.040%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	126110	36.512	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	146.040%#
66) 1,2-Dichlorobenzene-d4	13.848	152	43344	33.368	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	133.480%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW052424\
Data File : VW028799.D
Acq On : 25 May 2024 20:04
Operator : SY/MD
Sample : P2590-01
Misc : 4.88g/10mL/MSVOA_W/SOIL/A
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH6E7

Quant Time: May 27 02:24:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050224SMA.M
Quant Title : SFAM01.0
QLast Update : Mon May 27 02:22:53 2024
Response via : Initial Calibration

