

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_W\Data\VW052522\  
 Data File : VW023335.D  
 Acq On : 25 May 2022 21:52  
 Operator : SY/VA  
 Sample : N2995-01RE  
 Misc : 2.69g/10mL/MSVOA\_W/SOIL  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 GBGQ4

Quant Time: May 26 05:15:30 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_W\Method\SFAMWLM050222SMA.M  
 Quant Title : SFAM01.0  
 QLast Update : Thu May 26 04:46:06 2022  
 Response via : Initial Calibration


| Compound                      | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|--------|----------|
| -----                         |        |       |          |          |        |          |
| Internal Standards            |        |       |          |          |        |          |
| 1) 1,4-Difluorobenzene        | 8.842  | 114   | 317711   | 25.000   | ug/L   | 0.00     |
| 28) Chlorobenzene-d5          | 11.628 | 117   | 320053   | 25.000   | ug/L   | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 13.560 | 152   | 155261   | 25.000   | ug/L   | 0.00     |
|                               |        |       |          |          |        |          |
| System Monitoring Compounds   |        |       |          |          |        |          |
| 4) Vinyl Chloride-d3          | 2.349  | 65    | 98349    | 15.088   | ug/L   | 0.00     |
| Spiked Amount                 | 25.000 | Range | 30 - 150 | Recovery | =      | 60.360%  |
| 7) Chloroethane-d5            | 2.885  | 69    | 66584    | 15.701   | ug/L   | 0.00     |
| Spiked Amount                 | 25.000 | Range | 30 - 150 | Recovery | =      | 62.800%  |
| 11) 1,1-Dichloroethene-d2     | 4.019  | 63    | 99113    | 11.768   | ug/L   | 0.00     |
| Spiked Amount                 | 25.000 | Range | 45 - 110 | Recovery | =      | 47.080%  |
| 21) 2-Butanone-d5             | 7.074  | 46    | 71234    | 57.477   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000 | Range | 20 - 135 | Recovery | =      | 114.960% |
| 24) Chloroform-d              | 7.647  | 84    | 209174   | 22.472   | ug/L   | 0.00     |
| Spiked Amount                 | 25.000 | Range | 40 - 150 | Recovery | =      | 89.880%  |
| 26) 1,2-Dichloroethane-d4     | 8.305  | 65    | 122303   | 23.957   | ug/L   | 0.00     |
| Spiked Amount                 | 25.000 | Range | 70 - 130 | Recovery | =      | 95.840%  |
| 32) Benzene-d6                | 8.275  | 84    | 363403   | 19.699   | ug/L   | 0.00     |
| Spiked Amount                 | 25.000 | Range | 20 - 135 | Recovery | =      | 78.800%  |
| 36) 1,2-Dichloropropane-d6    | 9.274  | 67    | 120600   | 23.273   | ug/L   | 0.00     |
| Spiked Amount                 | 25.000 | Range | 70 - 120 | Recovery | =      | 93.080%  |
| 41) Toluene-d8                | 10.323 | 98    | 333736   | 18.728   | ug/L   | 0.00     |
| Spiked Amount                 | 25.000 | Range | 30 - 130 | Recovery | =      | 74.920%  |
| 43) trans-1,3-Dichloroprop... | 10.579 | 79    | 47853    | 20.056   | ug/L   | 0.00     |
| Spiked Amount                 | 25.000 | Range | 30 - 135 | Recovery | =      | 80.240%  |
| 47) 2-Hexanone-d5             | 10.927 | 63    | 56825    | 57.443   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000 | Range | 20 - 135 | Recovery | =      | 114.880% |
| 56) 1,1,2,2-Tetrachloroeth... | 12.688 | 84    | 131537   | 27.233   | ug/L   | 0.00     |
| Spiked Amount                 | 25.000 | Range | 45 - 120 | Recovery | =      | 108.920% |
| 66) 1,2-Dichlorobenzene-d4    | 13.853 | 152   | 123449   | 22.746   | ug/L   | 0.00     |
| Spiked Amount                 | 25.000 | Range | 75 - 120 | Recovery | =      | 91.000%  |
|                               |        |       |          |          |        |          |
| Target Compounds              |        |       |          |          | Qvalue |          |
| 13) Acetone                   | 4.123  | 43    | 11153    | 12.020   | ug/L   | 100      |
| -----                         |        |       |          |          |        |          |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_W\Data\VW052522\  
Data File : VW023335.D  
Acq On : 25 May 2022 21:52  
Operator : SY/VA  
Sample : N2995-01RE  
Misc : 2.69g/10mL/MSVOA\_W/SOIL  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_W  
ClientSampleId :  
GBGQ4

Quant Time: May 26 05:15:30 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_W\Method\SFAMWLM050222SMA.M  
Quant Title : SFAM01.0  
QLast Update : Thu May 26 04:46:06 2022  
Response via : Initial Calibration





#13  
 Acetone  
 Concen: 12.020 ug/L  
 RT: 4.123 min Scan# 30  
 Delta R.T. 0.006 min  
 Lab File: VW023335.D  
 Acq: 25 May 2022 21:52

Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 GBGQ4

Tgt Ion: 43 Resp: 11153  
 Ion Ratio Lower Upper  
 43 100  
 58 32.1 0.0 64.0

