

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW052720\
 Data File : VW015506.D
 Acq On : 27 May 2020 15:58
 Operator : SY/VA
 Sample : VIBLK
 Misc : 5.00G/10.0mL/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK

Quant Time: May 28 03:15:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM052620S.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 28 01:49:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	337820	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	306626	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	141957	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	63748	20.59	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.36%
7) Chloroethane-d5	2.89	69	64846	21.06	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.24%
10) 1,1-Dichloroethene-d2	4.01	63	102937	13.65	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	54.60%
20) 2-Butanone-d5	7.08	46	57488	56.88	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	113.76%
24) Chloroform-d	7.65	84	167775	21.48	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.92%
26) 1,2-Dichloroethane-d4	8.31	65	102692	23.63	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.52%
29) Benzene-d6	8.27	84	325684	21.66	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	86.64%
33) 1,2-Dichloropropane-d6	9.27	67	101570	22.23	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	88.92%
37) Toluene-d8	10.32	98	303336	20.75	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	83.00%
38) trans-1,3-Dichloropropene-	10.58	79	44175	20.90	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.60%
39) 2-Hexanone-d5	10.93	63	43411	54.87	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	109.74%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	88336	25.67	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	102.68%
61) 1,2-Dichlorobenzene-d4	13.85	152	110402	23.21	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	92.84%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW052720\
Data File : VW015506.D
Acq On : 27 May 2020 15:58
Operator : SY/VA
Sample : VIBLK
Misc : 5.00G/10.0mL/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK

Quant Time: May 28 03:15:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM052620S.M
Quant Title : VOC Analysis
QLast Update : Thu May 28 01:49:49 2020
Response via : Initial Calibration

