


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW052819\
 Data File : VW010559.D
 Acq On : 28 May 2019 09:54
 Operator : SY/VA
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W052019S.M
 Title : SW846 8260
 Last Update : Tue May 21 06:38:43 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.1	17202	PASS
75	95	30	60	46.9	53258	PASS
95	95	100	100	100.0	113624	PASS
96	95	5	9	6.6	7551	PASS
173	174	0.00	2	1.3	1166	PASS
174	95	50	100	79.0	89816	PASS
175	174	5	9	7.4	6680	PASS
176	174	95	101	96.2	86373	PASS
177	176	5	9	6.7	5784	PASS