

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW052919\  
 Data File : VW010593.D  
 Acq On : 29 May 2019 16:14  
 Operator : SY/VA  
 Sample : K3075-07  
 Misc : 6.02G/5ML/MSVOA W/SOIL  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 TEST-PIT-4

Quant Time: May 30 07:51:16 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_W\METHOD\82W052019S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue May 21 06:38:43 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|--------|---------|----------|
| 1) Pentafluorobenzene       | 7.95  | 168  | 120210   | 50.00  | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.84  | 114  | 217005   | 50.00  | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 11.63 | 117  | 190814   | 50.00  | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.56 | 152  | 79805    | 50.00  | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |        |         |          |
| 33) 1,2-Dichloroethane-d4   | 8.30  | 65   | 67726    | 54.58  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 109.16% |          |
| 35) Dibromofluoromethane    | 7.88  | 113  | 65238    | 49.54  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 99.08%  |          |
| 50) Toluene-d8              | 10.32 | 98   | 262831   | 52.25  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 104.50% |          |
| 62) 4-Bromofluorobenzene    | 12.62 | 95   | 95791    | 47.81  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 95.62%  |          |
| Target Compounds            |       |      |          |        |         |          |
| 16) Acetone                 | 4.14  | 43   | 8643     | 34.126 | ug/l    | 99       |
| 20) Methylene Chloride      | 4.92  | 84   | 9647     | 1.581  | ug/l    | 95       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW052919\  
Data File : VW010593.D  
Acq On : 29 May 2019 16:14  
Operator : SY/VA  
Sample : K3075-07  
Misc : 6.02G/5ML/MSVOA W/SOIL  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
MSVOA\_W  
ClientSampleId :  
TEST-PIT-4

Quant Time: May 30 07:51:16 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_W\METHOD\82W052019S.M  
Quant Title : SW846 8260  
QLast Update : Tue May 21 06:38:43 2019  
Response via : Initial Calibration

