


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW060319\
Data File : VW010651.D
Acq On : 03 Jun 2019 11:03
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W052019S.M
Title : SW846 8260
Last Update : Tue May 21 06:38:43 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.1	28165	PASS
75	95	30	60	45.7	85069	PASS
95	95	100	100	100.0	186069	PASS
96	95	5	9	6.5	12166	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	76.9	143104	PASS
175	174	5	9	7.8	11214	PASS
176	174	95	101	96.3	137877	PASS
177	176	5	9	6.6	9128	PASS