

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060421\
 Data File : VW019129.D
 Acq On : 04 Jun 2021 13:41
 Operator : SY/VA
 Sample : VIBLK124
 Misc : 5.00G/10ML/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK124

Quant Time: Jun 05 00:39:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060221SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Jun 05 00:35:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	917228	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.633	117	810797	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	400589	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	243008	20.036	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.160%
7) Chloroethane-d5	2.897	69	215026	20.815	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.280%
11) 1,1-Dichloroethene-d2	4.025	63	366683	15.216	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.880%
21) 2-Butanone-d5	7.079	46	123656	38.922	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.840%
24) Chloroform-d	7.652	84	550658	22.330	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.320%
26) 1,2-Dichloroethane-d4	8.305	65	298839	21.604	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	86.400%
32) Benzene-d6	8.274	84	1144499	24.776	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	99.120%
36) 1,2-Dichloropropane-d6	9.274	67	325229	23.224	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.880%
41) Toluene-d8	10.323	98	1082662	25.520	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	102.080%
43) trans-1,3-Dichloroprop...	10.579	79	141878	23.392	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	93.560%
47) 2-Hexanone-d5	10.926	63	105302	44.342	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.680%
56) 1,1,2,2-Tetrachloroeth...	12.694	84	235929	20.423	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	81.680%
66) 1,2-Dichlorobenzene-d4	13.853	152	356388	25.489	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	101.960%
Target Compounds						
					Qvalue	
16) Methylene chloride	4.927	84	31388	1.844	ug/L	94
60) Isopropylbenzene	12.463	105	22728	0.379	ug/L	96
63) 1,2,4-Trimethylbenzene	13.249	105	27374	0.557	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW060421\
Data File : VW019129.D
Acq On : 04 Jun 2021 13:41
Operator : SY/VA
Sample : VIBLK124
Misc : 5.00G/10ML/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK124

Quant Time: Jun 05 00:39:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060221SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Jun 05 00:35:45 2021
Response via : Initial Calibration

