

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW060519\
 Data File : VW010688.D
 Acq On : 05 Jun 2019 13:21
 Operator : SY/VA
 Sample : K3182-01
 Misc : 5.08G/5ML/MSVOA W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 B-06-2.5-3.0

Quant Time: Jun 06 07:41:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 05 03:21:19 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	154529	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	264583	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	229994	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	100020	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	103533	56.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.68%	
35) Dibromofluoromethane	7.88	113	78067	54.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.72%	
50) Toluene-d8	10.32	98	340687	54.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.20%	
62) 4-Bromofluorobenzene	12.62	95	117943	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	
Target Compounds						
16) Acetone	4.13	43	13240	27.129	ug/l	92
20) Methylene Chloride	4.92	84	3408	1.991	ug/l	95
95) Naphthalene	15.36	128	4462	1.230	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW060519\
Data File : VW010688.D
Acq On : 05 Jun 2019 13:21
Operator : SY/VA
Sample : K3182-01
Misc : 5.08G/5ML/MSVOA W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
B-06-2.5-3.0

Quant Time: Jun 06 07:41:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 05 03:21:19 2019
Response via : Initial Calibration

