

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW060519\
 Data File : VW010696.D
 Acq On : 05 Jun 2019 17:00
 Operator : SY/VA
 Sample : K3172-04
 Misc : 3.81G/5ML/MSVOA W/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 FK-SF-01-005-B

Quant Time: Jun 06 07:58:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 05 03:21:19 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	210778	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	373616	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	302076	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	108987	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	149334	60.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.22%	
35) Dibromofluoromethane	7.88	113	54682	26.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	53.92%	
50) Toluene-d8	10.32	98	465955	52.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.80%	
62) 4-Bromofluorobenzene	12.62	95	147281	45.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.02%	
Target Compounds						
16) Acetone	4.12	43	17971	26.996	ug/l	95
20) Methylene Chloride	4.92	84	9894	4.237	ug/l	98
68) m/p-Xylenes	11.84	106	8064	1.973	ug/l	96
69) o-Xylene	12.17	106	4975	1.284	ug/l	97
70) Styrene	12.18	104	53627	7.965	ug/l	92
80) 1,3,5-Trimethylbenzene	12.94	105	9700	1.447	ug/l	97
84) 1,2,4-Trimethylbenzene	13.25	105	28875	4.280	ug/l	97
95) Naphthalene	15.37	128	193494	48.949	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW060519\
Data File : VW010696.D
Acq On : 05 Jun 2019 17:00
Operator : SY/VA
Sample : K3172-04
Misc : 3.81G/5ML/MSVOA W/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
FK-SF-01-005-B

Quant Time: Jun 06 07:58:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 05 03:21:19 2019
Response via : Initial Calibration

