

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW060619\
 Data File : VW010705.D
 Acq On : 06 Jun 2019 13:04
 Operator : SY/VA
 Sample : K3222-02
 Misc : 4.87G/5ML/MSVOA W/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TT-007-IDW/ISO-20190605

Quant Time: Jun 07 08:17:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 05 03:21:19 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	209947	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	361062	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	308247	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	132262	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	144213	58.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.56%	
35) Dibromofluoromethane	7.88	113	106169	54.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.34%	
50) Toluene-d8	10.32	98	470081	54.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.40%	
62) 4-Bromofluorobenzene	12.62	95	160040	51.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.36%	
Target Compounds						
16) Acetone	4.13	43	13581	20.482	ug/l	99
18) Methyl Acetate	4.68	43	9004	6.308	ug/l	97
52) Toluene	10.39	92	23729	3.825	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW060619\
Data File : VW010705.D
Acq On : 06 Jun 2019 13:04
Operator : SY/VA
Sample : K3222-02
Misc : 4.87G/5ML/MSVOA W/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TT-007-IDW/ISO-20190605

Quant Time: Jun 07 08:17:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 05 03:21:19 2019
Response via : Initial Calibration

