

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW060619\
 Data File : VW010706.D
 Acq On : 06 Jun 2019 13:31
 Operator : SY/VA
 Sample : K3222-03
 Misc : 4.96G/5ML/MSVOA W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TT-008-IDW/ISO-20190605

Quant Time: Jun 07 08:19:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 05 03:21:19 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	220369	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	382377	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	338657	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	151210	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	153378	59.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.10%	
35) Dibromofluoromethane	7.88	113	113318	54.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.20%	
50) Toluene-d8	10.32	98	501420	55.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.18%	
62) 4-Bromofluorobenzene	12.62	95	175493	52.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.98%	
Target Compounds						
16) Acetone	4.12	43	43990	63.206	ug/l	Qvalue 100
18) Methyl Acetate	4.68	43	5422	3.619	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW060619\
Data File : VW010706.D
Acq On : 06 Jun 2019 13:31
Operator : SY/VA
Sample : K3222-03
Misc : 4.96G/5ML/MSVOA W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TT-008-IDW/ISO-20190605

Quant Time: Jun 07 08:19:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 05 03:21:19 2019
Response via : Initial Calibration

