


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW060619\
Data File : VW010700.D
Acq On : 06 Jun 2019 09:38
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
Title : SW846 8260
Last Update : Wed Jun 05 03:21:19 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.6	22442	PASS
75	95	30	60	53.6	51000	PASS
95	95	100	100	100.0	95096	PASS
96	95	5	9	6.8	6453	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	74.7	71042	PASS
175	174	5	9	7.7	5495	PASS
176	174	95	101	96.5	68584	PASS
177	176	5	9	6.9	4702	PASS