

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW060719\
 Data File : VW010728.D
 Acq On : 07 Jun 2019 14:52
 Operator : SY/VA
 Sample : K3247-04
 Misc : 3.95G/5ML/MSVOA W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 FK-GF-02-041-B

Quant Time: Jun 08 03:48:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 05 03:21:19 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	209985	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	360846	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	310885	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	129361	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	145993	58.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.98%	
35) Dibromofluoromethane	7.88	113	104759	53.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.98%	
50) Toluene-d8	10.32	98	468700	54.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.14%	
62) 4-Bromofluorobenzene	12.62	95	160007	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
Target Compounds						
16) Acetone	4.12	43	43987	66.327	ug/l	98
20) Methylene Chloride	4.92	84	8372	3.599	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW060719\
Data File : VW010728.D
Acq On : 07 Jun 2019 14:52
Operator : SY/VA
Sample : K3247-04
Misc : 3.95G/5ML/MSVOA W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
FK-GF-02-041-B

Quant Time: Jun 08 03:48:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 05 03:21:19 2019
Response via : Initial Calibration

