

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW060719\
 Data File : VW010723.D
 Acq On : 07 Jun 2019 10:58
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 08 04:09:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 05 03:21:19 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.462	0.441	4.5	105	0.00
3 P	Chloromethane	0.680	0.663	2.5	104	0.00
4 C	Vinyl Chloride	0.628	0.617	1.8#	103	0.00
5 T	Bromomethane	0.345	0.334	3.2	104	0.00
6 T	Chloroethane	0.377	0.376	0.3	100	0.00
7 T	Trichlorofluoromethane	0.815	0.852	-4.5	105	0.00
8 T	Diethyl Ether	0.291	0.274	5.8	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.501	0.525	-4.8	104	0.00
10 T	Methyl Iodide	0.443	0.503	-13.5	101	0.00
11 T	Tert butyl alcohol	0.048	0.044	8.3	97	0.00
12 CM	1,1-Dichloroethene	0.484	0.501	-3.5#	104	0.00
13 T	Acrolein	0.020	0.023	-15.0	108	0.00
14 T	Allyl chloride	1.066	1.135	-6.5	105	0.00
15 T	Acrylonitrile	0.142	0.125	12.0	91	0.00
16 T	Acetone	0.158	0.160	-1.3	96	0.00
17 T	Carbon Disulfide	1.535	1.590	-3.6	105	0.00
18 T	Methyl Acetate	0.340	0.287	15.6	91	0.00
19 T	Methyl tert-butyl Ether	1.395	1.284	8.0	92	0.00
20 T	Methylene Chloride	0.554	0.585	-5.6	109	0.00
21 T	trans-1,2-Dichloroethene	0.549	0.563	-2.6	101	0.00
22 T	Diisopropyl ether	2.134	2.154	-0.9	99	0.00
23 T	Vinyl Acetate	1.288	1.248	3.1	95	0.00
24 P	1,1-Dichloroethane	1.070	1.110	-3.7	104	0.00
25 T	2-Butanone	0.219	0.204	6.8	91	0.00
26 T	2,2-Dichloropropane	0.870	0.982	-12.9	112	0.00
27 T	cis-1,2-Dichloroethene	0.604	0.614	-1.7	101	0.00
28 T	Bromochloromethane	0.477	0.471	1.3	95	0.00
29 T	Tetrahydrofuran	0.138	0.116	15.9	87	0.00
30 C	Chloroform	0.978	1.010	-3.3#	103	0.00
31 T	Cyclohexane	1.189	1.179	0.8	104	0.00
32 T	1,1,1-Trichloroethane	0.855	0.911	-6.5	106	0.00
33 S	1,2-Dichloroethane-d4	0.589	0.538	8.7	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.271	0.269	0.7	89	0.00
36 T	1,1-Dichloropropene	0.512	0.536	-4.7	104	0.00
37 T	Ethyl Acetate	0.284	0.254	10.6	91	0.00
38 T	Carbon Tetrachloride	0.442	0.480	-8.6	105	0.00
39 T	Methylcyclohexane	0.651	0.678	-4.1	102	0.00
40 TM	Benzene	1.396	1.423	-1.9	101	0.00
41 T	Methacrylonitrile	0.169	0.153	9.5	91	0.00
42 TM	1,2-Dichloroethane	0.443	0.435	1.8	97	0.00
43 T	Isopropyl Acetate	0.548	0.490	10.6	89	0.00
44 TM	Trichloroethene	0.344	0.356	-3.5	103	0.00
45 C	1,2-Dichloropropane	0.370	0.374	-1.1#	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW060719\
 Data File : VW010723.D
 Acq On : 07 Jun 2019 10:58
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 08 04:09:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 05 03:21:19 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.172	0.163	5.2	94	0.00
47 T	Bromodichloromethane	0.445	0.455	-2.2	100	0.00
48 T	Methyl methacrylate	0.264	0.255	3.4	95	0.00
49 T	1,4-Dioxane	0.003	0.002	33.3#	94	0.00
50 S	Toluene-d8	1.190	1.135	4.6	90	0.00
51 T	4-Methyl-2-Pentanone	0.280	0.247	11.8	89	0.00
52 CM	Toluene	0.859	0.885	-3.0#	101	0.00
53 T	t-1,3-Dichloropropene	0.471	0.477	-1.3	98	0.00
54 T	cis-1,3-Dichloropropene	0.555	0.575	-3.6	100	0.00
55 T	1,1,2-Trichloroethane	0.247	0.233	5.7	92	0.00
56 T	Ethyl methacrylate	0.381	0.352	7.6	91	0.00
57 T	1,3-Dichloropropane	0.466	0.441	5.4	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.199	0.183	8.0	96	0.00
59 T	2-Hexanone	0.202	0.185	8.4	91	0.00
60 T	Dibromochloromethane	0.274	0.271	1.1	96	0.00
61 T	1,2-Dibromoethane	0.234	0.220	6.0	93	0.00
62 S	4-Bromofluorobenzene	0.433	0.409	5.5	90	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.332	0.333	-0.3	97	0.00
65 PM	Chlorobenzene	0.983	0.998	-1.5	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.340	0.353	-3.8	99	0.00
67 C	Ethyl Benzene	1.890	1.980	-4.8#	103	0.00
68 T	m/p-Xylenes	0.676	0.705	-4.3	102	0.00
69 T	o-Xylene	0.642	0.667	-3.9	101	0.00
70 T	Styrene	1.114	1.150	-3.2	100	0.00
71 P	Bromoform	0.176	0.165	6.2	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.757	4.025	-7.1	102	0.00
74 T	N-amyl acetate	1.214	1.150	5.3	92	0.00
75 P	1,1,2,2-Tetrachloroethane	0.715	0.657	8.1	90	0.00
76 T	1,2,3-Trichloropropane	0.525	0.492	6.3	93	0.00
77 T	Bromobenzene	0.820	0.844	-2.9	98	0.00
78 T	n-propylbenzene	4.531	4.888	-7.9	103	0.00
79 T	2-Chlorotoluene	2.588	2.748	-6.2	102	0.00
80 T	1,3,5-Trimethylbenzene	3.075	3.297	-7.2	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.228	0.220	3.5	91	0.00
82 T	4-Chlorotoluene	2.671	2.865	-7.3	104	0.00
83 T	tert-Butylbenzene	2.610	2.759	-5.7	101	0.00
84 T	1,2,4-Trimethylbenzene	3.095	3.276	-5.8	102	0.00
85 T	sec-Butylbenzene	3.800	4.070	-7.1	102	0.00
86 T	p-Isopropyltoluene	3.398	3.652	-7.5	103	0.00
87 T	1,3-Dichlorobenzene	1.582	1.646	-4.0	99	0.00
88 T	1,4-Dichlorobenzene	1.568	1.603	-2.2	100	0.00
89 T	n-Butylbenzene	3.391	3.729	-10.0	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW060719\
Data File : VW010723.D
Acq On : 07 Jun 2019 10:58
Operator : SY/VA
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Jun 08 04:09:53 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 05 03:21:19 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.609	0.681	-11.8	104	0.00
91 T	1,2-Dichlorobenzene	1.422	1.446	-1.7	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.123	0.116	5.7	94	0.00
93 T	1,2,4-Trichlorobenzene	0.998	1.066	-6.8	101	0.00
94 T	Hexachlorobutadiene	0.626	0.676	-8.0	103	0.00
95 T	Naphthalene	1.814	1.780	1.9	93	0.00
96 T	1,2,3-Trichlorobenzene	0.849	0.892	-5.1	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6