

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW060719\
 Data File : VW010734.D
 Acq On : 07 Jun 2019 17:35
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 08 04:03:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 05 03:21:19 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	0.462	0.433	6.3	95	0.00
3 P	Chloromethane	0.680	0.655	3.7	95	0.00
4 C	Vinyl Chloride	0.628	0.599	4.6#	92	0.00
5 T	Bromomethane	0.345	0.354	-2.6	102	0.00
6 T	Chloroethane	0.377	0.373	1.1	91	0.00
7 T	Trichlorofluoromethane	0.815	0.813	0.2	92	0.00
8 T	Diethyl Ether	0.291	0.281	3.4	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.501	0.502	-0.2	91	0.01
10 T	Methyl Iodide	0.443	0.448	-1.1	83	0.00
11 T	Tert butyl alcohol	0.048	0.045	6.3	91	0.00
12 CM	1,1-Dichloroethene	0.484	0.475	1.9#	91	0.00
13 T	Acrolein	0.020	0.012	40.0#	54	0.00
14 T	Allyl chloride	1.066	1.086	-1.9	93	0.00
15 T	Acrylonitrile	0.142	0.137	3.5	92	0.00
16 T	Acetone	0.158	0.155	1.9	86	0.00
17 T	Carbon Disulfide	1.535	1.497	2.5	91	0.00
18 T	Methyl Acetate	0.340	0.314	7.6	92	0.00
19 T	Methyl tert-butyl Ether	1.395	1.347	3.4	89	0.00
20 T	Methylene Chloride	0.554	0.549	0.9	94	0.00
21 T	trans-1,2-Dichloroethene	0.549	0.542	1.3	89	0.00
22 T	Diisopropyl ether	2.134	2.182	-2.2	93	0.00
23 T	Vinyl Acetate	1.288	1.293	-0.4	90	0.00
24 P	1,1-Dichloroethane	1.070	1.093	-2.1	94	0.00
25 T	2-Butanone	0.219	0.210	4.1	87	0.00
26 T	2,2-Dichloropropane	0.870	0.916	-5.3	96	0.00
27 T	cis-1,2-Dichloroethene	0.604	0.603	0.2	91	0.00
28 T	Bromochloromethane	0.477	0.485	-1.7	90	0.00
29 T	Tetrahydrofuran	0.138	0.128	7.2	89	0.00
30 C	Chloroform	0.978	0.989	-1.1#	93	0.00
31 T	Cyclohexane	1.189	1.141	4.0	93	0.00
32 T	1,1,1-Trichloroethane	0.855	0.880	-2.9	94	0.00
33 S	1,2-Dichloroethane-d4	0.589	0.596	-1.2	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	87	0.00
35 S	Dibromofluoromethane	0.271	0.287	-5.9	88	0.00
36 T	1,1-Dichloropropene	0.512	0.514	-0.4	92	0.00
37 T	Ethyl Acetate	0.284	0.270	4.9	89	0.00
38 T	Carbon Tetrachloride	0.442	0.447	-1.1	90	0.00
39 T	Methylcyclohexane	0.651	0.650	0.2	90	0.00
40 TM	Benzene	1.396	1.403	-0.5	92	0.00
41 T	Methacrylonitrile	0.169	0.177	-4.7	97	0.00
42 TM	1,2-Dichloroethane	0.443	0.446	-0.7	92	0.00
43 T	Isopropyl Acetate	0.548	0.527	3.8	89	0.00
44 TM	Trichloroethene	0.344	0.342	0.6	91	0.00
45 C	1,2-Dichloropropane	0.370	0.375	-1.4#	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW060719\
 Data File : VW010734.D
 Acq On : 07 Jun 2019 17:35
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 08 04:03:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 05 03:21:19 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.172	0.171	0.6	91	0.00
47 T	Bromodichloromethane	0.445	0.453	-1.8	92	0.00
48 T	Methyl methacrylate	0.264	0.259	1.9	89	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	93	0.00
50 S	Toluene-d8	1.190	1.188	0.2	87	0.00
51 T	4-Methyl-2-Pentanone	0.280	0.269	3.9	90	0.00
52 CM	Toluene	0.859	0.859	0.0	90	0.00
53 T	t-1,3-Dichloropropene	0.471	0.473	-0.4	90	0.00
54 T	cis-1,3-Dichloropropene	0.555	0.565	-1.8	91	0.00
55 T	1,1,2-Trichloroethane	0.247	0.243	1.6	89	0.00
56 T	Ethyl methacrylate	0.381	0.373	2.1	89	0.00
57 T	1,3-Dichloropropane	0.466	0.460	1.3	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.199	0.203	-2.0	98	0.00
59 T	2-Hexanone	0.202	0.196	3.0	89	0.00
60 T	Dibromochloromethane	0.274	0.276	-0.7	90	0.00
61 T	1,2-Dibromoethane	0.234	0.227	3.0	88	0.00
62 S	4-Bromofluorobenzene	0.433	0.444	-2.5	90	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.332	0.322	3.0	87	0.00
65 PM	Chlorobenzene	0.983	0.981	0.2	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.340	0.343	-0.9	90	0.00
67 C	Ethyl Benzene	1.890	1.907	-0.9#	92	0.00
68 T	m/p-Xylenes	0.676	0.689	-1.9	93	0.00
69 T	o-Xylene	0.642	0.646	-0.6	92	0.00
70 T	Styrene	1.114	1.132	-1.6	92	0.00
71 P	Bromoform	0.176	0.171	2.8	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.757	3.747	0.3	93	0.00
74 T	N-amyl acetate	1.214	1.182	2.6	93	0.00
75 P	1,1,2,2-Tetrachloroethane	0.715	0.672	6.0	89	0.00
76 T	1,2,3-Trichloropropane	0.525	0.493	6.1	91	0.00
77 T	Bromobenzene	0.820	0.813	0.9	93	0.00
78 T	n-propylbenzene	4.531	4.583	-1.1	94	0.00
79 T	2-Chlorotoluene	2.588	2.583	0.2	94	0.00
80 T	1,3,5-Trimethylbenzene	3.075	3.158	-2.7	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.228	0.214	6.1	86	0.00
82 T	4-Chlorotoluene	2.671	2.692	-0.8	95	0.00
83 T	tert-Butylbenzene	2.610	2.655	-1.7	95	0.00
84 T	1,2,4-Trimethylbenzene	3.095	3.152	-1.8	96	0.00
85 T	sec-Butylbenzene	3.800	3.851	-1.3	95	0.00
86 T	p-Isopropyltoluene	3.398	3.482	-2.5	96	0.00
87 T	1,3-Dichlorobenzene	1.582	1.587	-0.3	93	0.00
88 T	1,4-Dichlorobenzene	1.568	1.562	0.4	95	0.00
89 T	n-Butylbenzene	3.391	3.562	-5.0	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW060719\
Data File : VW010734.D
Acq On : 07 Jun 2019 17:35
Operator : SY/VA
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Jun 08 04:03:55 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 05 03:21:19 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.609	0.619	-1.6	93	0.00
91 T	1,2-Dichlorobenzene	1.422	1.409	0.9	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.123	0.117	4.9	93	0.00
93 T	1,2,4-Trichlorobenzene	0.998	1.036	-3.8	96	0.00
94 T	Hexachlorobutadiene	0.626	0.650	-3.8	97	0.00
95 T	Naphthalene	1.814	1.807	0.4	92	0.00
96 T	1,2,3-Trichlorobenzene	0.849	0.880	-3.7	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5