


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW060920\
 Data File : VW015564.D
 Acq On : 09 Jun 2020 10:35
 Operator : SY/VA
 Sample : BFB07
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB07

Integration File: LSCINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM052620S.M
 Title : VOC Analysis
 Last Update : Tue Jun 09 12:46:40 2020



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.6	16504	PASS
75	95	30	80	53.3	40650	PASS
95	95	100	100	100.0	76312	PASS
96	95	5	9	7.0	5340	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	75.5	57618	PASS
175	174	5	9	8.2	4738	PASS
176	174	95	101	97.2	55989	PASS
177	176	5	9	7.0	3911	PASS

m/z	Abundance
36.00	835.0
37.00	4585.0
38.00	3981.0
39.00	1570.0
39.90	242.0
40.80	53.0
42.90	175.0
44.00	824.0
45.00	1290.0
45.90	105.0
47.00	1059.0
47.90	560.0
49.00	3885.0
50.00	17472.0
51.00	5346.0
51.90	285.0
54.90	291.0
55.90	1499.0
57.00	2658.0
57.90	84.0
59.00	427.0
60.00	917.0
61.00	4297.0
62.00	4345.0
63.00	3324.0
63.90	308.0
65.10	253.0
66.80	288.0
68.00	8293.0
69.00	9132.0
69.90	577.0
71.00	87.0
72.00	521.0
73.00	5968.0
74.00	14460.0
75.00	41560.0
76.00	3684.0
77.00	487.0
77.90	396.0
78.90	2604.0
79.90	775.0
80.90	2869.0
81.90	744.0
82.90	149.0
86.90	3012.0
87.90	2854.0
88.90	270.0
90.90	349.0
91.90	2298.0
93.00	3263.0
94.00	9164.0
95.00	75048.0
96.00	5600.0
97.20	173.0
102.90	512.0
103.90	543.0
105.00	264.0
105.90	393.0
106.90	108.0
109.90	185.0
111.10	316.0
111.80	184.0
112.70	64.0
115.00	319.0
115.90	388.0
116.90	633.0
118.00	548.0
118.90	818.0
120.80	101.0
125.00	824.0
125.70	229.0
126.60	78.0
127.90	269.0
128.80	134.0
129.80	234.0
130.90	175.0
132.90	2569.0
133.70	586.0
134.90	272.0
135.90	57.0
137.00	164.0
139.70	58.0
140.90	705.0
141.80	94.0
142.90	678.0
144.90	102.0
145.90	157.0
146.90	349.0
147.90	168.0
148.90	216.0
149.70	51.0
153.60	52.0

Instrument :
MSVOA_W
ClientSampleId :
BFB07

154.70	177.0
156.90	99.0
159.00	142.0
160.90	277.0
161.60	58.0
163.00	488.0
163.90	115.0
164.90	232.0
166.00	51.0
172.00	372.0
173.90	51192.0
174.90	4759.0
175.90	51216.0
176.90	3771.0
177.90	309.0
178.90	556.0
179.90	119.0
188.90	282.0
191.00	2047.0
192.10	601.0
193.00	3050.0
193.90	533.0
194.90	332.0
203.00	282.0
203.90	55.0
205.00	527.0
206.00	138.0
207.00	1739.0
208.00	337.0
209.00	244.0
218.90	58.0
220.80	82.0
234.90	312.0
235.90	55.0
248.00	52.0
248.90	1831.0
249.90	570.0
251.00	775.0
251.90	258.0
252.90	148.0
265.00	2201.0
266.00	561.0
266.80	407.0
267.90	115.0
277.30	51.0
278.40	155.0
279.10	244.0
281.00	30392.0
282.00	8443.0
283.00	4849.0
283.90	843.0
284.90	253.0

Instrument :
MSVOA_W
ClientSampleId :
BFB07

m/z	Abundance
36.00	936.0
37.00	4955.0
38.00	4273.0
39.00	1577.0
39.90	224.0
42.90	136.0
43.90	836.0
45.00	1131.0
46.10	152.0
47.00	1020.0
48.00	633.0
49.00	4359.0
50.00	17816.0
51.00	5503.0
52.10	226.0
55.00	260.0
56.10	1436.0
57.00	2845.0
58.00	131.0
59.00	221.0
60.00	1012.0
61.00	4741.0
62.00	4597.0
63.00	3636.0
64.10	384.0
65.10	221.0
66.80	234.0
68.00	9269.0
69.00	9435.0
70.00	727.0
71.00	79.0
72.00	482.0
73.00	5421.0
74.00	15865.0
75.00	45312.0
76.00	4030.0
77.00	457.0
78.10	336.0
78.90	2979.0
80.00	964.0
80.90	2895.0
81.90	753.0
83.00	91.0
85.90	103.0
86.90	3312.0
88.00	2976.0
91.00	354.0
92.00	2428.0
92.90	3797.0
94.00	10364.0
95.00	84664.0
96.00	5980.0
96.90	274.0
102.90	348.0
103.90	546.0
104.80	200.0
105.90	385.0
107.00	64.0
109.90	166.0
110.90	166.0
111.70	89.0
112.20	152.0
112.80	94.0
114.90	234.0
115.80	298.0
116.80	575.0
117.90	631.0
118.90	681.0
124.80	444.0
125.70	116.0
126.10	106.0
127.80	287.0
129.00	201.0
129.80	324.0
130.90	120.0
131.90	51.0
133.00	1672.0
134.80	232.0
135.90	54.0
137.00	144.0
139.90	86.0
140.90	767.0
141.80	117.0
142.80	827.0
143.90	53.0
144.80	129.0
145.80	130.0
146.90	184.0
148.00	218.0
148.90	122.0
149.80	142.0
152.90	64.0

Instrument :
MSVOA_W
ClientSampleId :
BFB07

154.80	157.0
157.00	192.0
158.90	91.0
161.00	204.0
161.60	53.0
163.00	326.0
163.80	90.0
164.80	214.0
171.00	90.0
171.90	333.0
173.90	65000.0
174.90	5174.0
176.00	59456.0
176.90	4372.0
177.80	311.0
178.80	379.0
179.80	57.0
188.90	197.0
190.90	1451.0
192.00	325.0
192.90	1971.0
193.90	488.0
194.90	229.0
202.90	180.0
203.80	59.0
204.90	323.0
205.90	121.0
207.00	1021.0
208.00	159.0
208.80	150.0
219.00	64.0
221.10	51.0
222.80	63.0
235.00	208.0
248.90	1194.0
249.80	353.0
250.90	503.0
252.00	115.0
253.10	60.0
264.90	1683.0
265.90	399.0
267.00	403.0
268.00	130.0
278.10	83.0
279.50	201.0
281.00	19032.0
282.00	4783.0
283.00	2954.0
284.00	723.0
285.10	179.0

Instrument :
MSVOA_W
ClientSampleId :
BFB07

m/z	Abundance
36.00	689.0
37.00	3665.0
38.00	3144.0
39.00	1320.0
39.90	187.0
43.00	99.0
44.00	687.0
45.00	788.0
45.90	103.0
46.90	876.0
47.90	557.0
49.00	2890.0
50.00	14226.0
51.00	4083.0
52.00	237.0
55.10	189.0
56.00	1152.0
57.00	1977.0
58.00	125.0
59.10	121.0
59.90	690.0
61.00	3354.0
62.00	3563.0
63.00	2544.0
63.90	298.0
64.80	143.0
66.90	159.0
67.90	6868.0
69.00	7521.0
70.00	540.0
72.00	391.0
73.00	3750.0
74.00	12741.0
75.00	35080.0
76.00	3113.0
77.00	397.0
78.00	245.0
78.90	2226.0
79.90	686.0
80.90	2443.0
81.90	564.0
83.00	94.0
86.10	117.0
86.90	2558.0
87.90	2603.0
89.00	65.0
90.90	361.0
91.90	2141.0
93.00	2893.0
94.00	8221.0
95.00	69224.0
96.00	4442.0
97.00	146.0
102.90	178.0
103.90	387.0
104.80	120.0
105.80	323.0
106.80	104.0
109.80	80.0
110.90	103.0
111.80	85.0
112.90	60.0
114.90	148.0
115.80	236.0
116.90	435.0
117.90	410.0
118.90	519.0
125.10	200.0
125.70	67.0
126.60	53.0
127.90	283.0
129.00	84.0
129.80	368.0
130.70	80.0
133.00	631.0
133.70	299.0
134.90	204.0
137.00	112.0
140.00	81.0
140.90	666.0
141.90	114.0
142.90	591.0
144.90	84.0
145.80	111.0
146.00	113.0
146.90	196.0
147.80	213.0
148.80	76.0
149.90	111.0
152.80	68.0
155.00	133.0
156.90	108.0

Instrument :
MSVOA_W
ClientSampleId :
BFB07

158.80	83.0
160.80	154.0
162.90	162.0
165.00	67.0
170.80	72.0
171.80	248.0
173.90	56664.0
174.90	4282.0
175.90	57296.0
176.90	3590.0
177.80	159.0
179.00	149.0
188.90	66.0
190.90	556.0
191.90	173.0
192.90	933.0
194.20	68.0
194.90	61.0
202.80	93.0
204.90	178.0
207.00	520.0
208.00	104.0
209.00	90.0
234.90	69.0
249.00	640.0
249.90	110.0
251.00	226.0
265.00	788.0
267.00	132.0
268.10	77.0
279.10	110.0
281.00	9646.0
282.00	2197.0
283.00	1441.0
283.90	391.0

Instrument :
MSVOA_W
ClientSampleId :
BFB07

Scan 1749 (12.567 min): VW015564.D
BFB07

m/z	Abundance
39.90	153.0
41.10	59.0
43.90	235.0

Instrument :
MSVOA_W
ClientSampleId :
BFB07