


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW061019\
Data File : VW010735.D
Acq On : 10 Jun 2019 10:09
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
Title : SW846 8260
Last Update : Wed Jun 05 03:21:19 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.9	23152	PASS
75	95	30	60	55.0	55672	PASS
95	95	100	100	100.0	101282	PASS
96	95	5	9	6.8	6838	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	74.7	75701	PASS
175	174	5	9	8.0	6057	PASS
176	174	95	101	96.5	73082	PASS
177	176	5	9	6.7	4898	PASS