

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW061219\
 Data File : VW010778.D
 Acq On : 12 Jun 2019 15:15
 Operator : SY/VA
 Sample : K3303-01
 Misc : 5.71G/5ML/MSVOA W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BAT-B03

Quant Time: Jun 13 08:07:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 05 03:21:19 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	196698	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	341445	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	296025	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	129399	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	135707	58.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.06%	
35) Dibromofluoromethane	7.88	113	99624	53.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.52%	
50) Toluene-d8	10.32	98	451646	55.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.14%	
62) 4-Bromofluorobenzene	12.62	95	154071	52.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.20%	
Target Compounds						
16) Acetone	4.13	43	8121	13.073	ug/l	94
20) Methylene Chloride	4.91	84	12414	5.696	ug/l	94
52) Toluene	10.39	92	41918	7.145	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW061219\
Data File : VW010778.D
Acq On : 12 Jun 2019 15:15
Operator : SY/VA
Sample : K3303-01
Misc : 5.71G/5ML/MSVOA W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BAT-B03

Quant Time: Jun 13 08:07:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 05 03:21:19 2019
Response via : Initial Calibration

