

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW061219\
 Data File : VW010783.D
 Acq On : 12 Jun 2019 17:35
 Operator : SY/VA
 Sample : K3303-03
 Misc : 5.02G/5ML/MSVOA W/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BAT-B02

Quant Time: Jun 13 08:16:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 05 03:21:19 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	192788	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	337168	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	285477	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	109713	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	135863	59.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.58%	
35) Dibromofluoromethane	7.88	113	100892	55.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.26%	
50) Toluene-d8	10.32	98	439660	54.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.58%	
62) 4-Bromofluorobenzene	12.62	95	143968	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
Target Compounds						
16) Acetone	4.13	43	12671	20.811	ug/l	96
20) Methylene Chloride	4.91	84	12472	5.839	ug/l	94
64) Tetrachloroethene	10.87	164	21281	11.227	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW061219\
Data File : VW010783.D
Acq On : 12 Jun 2019 17:35
Operator : SY/VA
Sample : K3303-03
Misc : 5.02G/5ML/MSVOA W/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BAT-B02

Quant Time: Jun 13 08:16:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 05 03:21:19 2019
Response via : Initial Calibration

