

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW061319\
 Data File : VW010794.D
 Acq On : 13 Jun 2019 13:35
 Operator : SY/VA
 Sample : K3343-02RE
 Misc : 4.67G/5ML/MSVOA W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Jun 14 07:56:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 05 03:21:19 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	192468	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	327685	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	288214	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	129436	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	108796	47.96	ug/l	0.00
Spiked Amount			Recovery	=	95.92%	
35) Dibromofluoromethane	7.88	113	1541	0.87	ug/l	0.00
Spiked Amount			Recovery	=	1.74%	
50) Toluene-d8	10.32	98	294974	37.82	ug/l	0.00
Spiked Amount			Recovery	=	75.64%	
62) 4-Bromofluorobenzene	12.62	95	114418	40.32	ug/l	0.00
Spiked Amount			Recovery	=	80.64%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Methyl Iodide	4.26	142	121	4.853	ug/l #	64
11) Tert butyl alcohol	5.12	59	1571	8.500	ug/l #	75
13) Acrolein	3.90	56	115	1.523	ug/l #	1
16) Acetone	4.12	43	25645	42.189	ug/l	97
17) Carbon Disulfide	4.38	76	9172	1.552	ug/l #	94
20) Methylene Chloride	4.92	84	10798	5.064	ug/l	94
25) 2-Butanone	7.18	43	5437	6.451	ug/l #	90
29) Tetrahydrofuran	7.54	42	789	1.488	ug/l #	81
31) Cyclohexane	7.95	56	5739	1.254	ug/l #	15
36) 1,1-Dichloropropene	7.95	75	17214	5.132	ug/l #	51
37) Ethyl Acetate	7.18	43	5437	2.925	ug/l #	72
38) Carbon Tetrachloride	7.95	117	19392	6.697	ug/l #	16
48) Methyl methacrylate	9.41	41	2939	1.698	ug/l #	44
51) 4-Methyl-2-Pentanone	10.21	43	5839	3.183	ug/l	99
59) 2-Hexanone	10.98	43	1808	1.364	ug/l	72
67) Ethyl Benzene	11.73	91	11717	1.076	ug/l	97
68) m/p-Xylenes	11.83	106	6716	1.722	ug/l	98
69) o-Xylene	12.16	106	8908	2.409	ug/l	91
74) N-amyl acetate	12.24	43	7286	2.318	ug/l #	52
76) 1,2,3-Trichloropropane	12.62	75	62493	45.976	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.94	105	20539	2.580	ug/l	97
81) trans-1,4-Dichloro-2-buten	12.62	75	62493	105.668	ug/l #	5
84) 1,2,4-Trimethylbenzene	13.26	105	64155	8.008	ug/l	99
85) sec-Butylbenzene	13.26	105	64155	6.521	ug/l #	57
90) Hexachloroethane	14.10	117	3460	2.196	ug/l #	11
92) 1,2-Dibromo-3-Chloropropan	14.46	75	946	2.975	ug/l #	6
95) Naphthalene	15.37	128	16210	3.453	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW061319\
Data File : VW010794.D
Acq On : 13 Jun 2019 13:35
Operator : SY/VA
Sample : K3343-02RE
Misc : 4.67G/5ML/MSVOA W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Jun 14 07:56:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 05 03:21:19 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.95 min Scan# 992

Delta R.T. 0.01 min

Lab File: VW010794.D

Acq: 13 Jun 2019 13:35

Instrument :

MSVOA_W


ClientSampleId :

Tot Ion:168 Resp: 192468

Ion Ratio Lower Upper




168 100

99 59.2 47.7 71.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VW



#10

Methyl Iodide

Concen: 4.853 ug/l

RT: 4.26 min Scan# 387

Delta R.T. -0.01 min

Lab File: VW010794.D

Acq: 13 Jun 2019 13:35


Tgt Ion:142 Resp: 121

Ion Ratio Lower Upper

142 100

127 19.8 34.6 51.8#



141 0.0 11.6 17.4#




Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V









#11

Tert butyl alcohol
 Concen: 8.500 ug/l
 RT: 5.12 min Scan# 528
 Delta R.T. -0.05 min
 Lab File: VW010794.D
 Acq: 13 Jun 2019 13:35

Instrument :
 MSVOA_W
 ClientSampled :




Tot Ion: 59 Resp: 1571
 Ion Ratio Lower Upper
 59 100
 57 1.3 8.6 12.8#




#13

Acrolein
 Concen: 1.523 ug/l
 RT: 3.90 min Scan# 328
 Delta R.T. 0.01 min
 Lab File: VW010794.D
 Acq: 13 Jun 2019 13:35

Tgt Ion: 56 Resp: 115
 Ion Ratio Lower Upper
 56 100
 55 238.3 57.6 86.4#





#16

Acetone

Concen: 42.189 ug/l

RT: 4.12 min Scan# 364

Delta R.T. -0.00 min

Lab File: VW010794.D

Acq: 13 Jun 2019 13:35

Instrument :

MSVOA_W


ClientSampled :

Tot Ion: 43 Resp: 25645

Ion Ratio Lower Upper

43 100

58 28.9 22.1 33.1



Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 58.00 (57.70 to 58.70): VW

4.12

10000

8000

6000



4000

2000

0

Time-->

4.00 4.10 4.20 4.30



#17

Carbon Disulfide

Concen: 1.552 ug/l

RT: 4.38 min Scan# 406

Delta R.T. 0.01 min

Lab File: VW010794.D


Acq: 13 Jun 2019 13:35

Tgt Ion: 76 Resp: 9172

Ion Ratio Lower Upper

76 100

78 7.2 7.5 11.3#



Abundance Ion 75.90 (75.60 to 76.60): VW

Ion 77.90 (77.60 to 78.60): VW

4.38

4000

3000


2000


1000

0

Time-->

4.25 4.30 4.35 4.40 4.45









#20
Methylene Chloride
Concen: 5.064 ug/l
RT: 4.92 min Scan# 494
Delta R.T. -0.00 min
Lab File: VW010794.D
Acq: 13 Jun 2019 13:35

Instrument :
MSVOA_W
ClientSampleId :

Tot Ion: 84 Resp: 10798
Ion Ratio Lower Upper
84 100
49 138.6 116.7 175.1
51 49.6 34.2 51.2
86 67.0 51.2 76.8




Abundance Ion 83.90 (83.60 to 84.60): VW
Ion 48.90 (48.60 to 49.60): VW
Ion 50.90 (50.60 to 51.60): VW
Ion 85.90 (85.60 to 86.60): VW






#25
2-Butanone
Concen: 6.451 ug/l
RT: 7.18 min Scan# 866
Delta R.T. 0.01 min
Lab File: VW010794.D
Acq: 13 Jun 2019 13:35

Tgt Ion: 43 Resp: 5437
Ion Ratio Lower Upper
43 100
72 17.9 18.1 27.1#



Abundance Ion 43.00 (42.70 to 43.70): VW
Ion 72.00 (71.70 to 72.70): VW









#29
Tetrahydrofuran
Concen: 1.488 ug/l
RT: 7.54 min Scan# 925
Delta R.T. 0.01 min
Lab File: VW010794.D
Acq: 13 Jun 2019 13:35

Instrument :
MSVOA_W
ClientSampled :

Tot Ion: 42 Resp: 789
Ion Ratio Lower Upper
42 100
72 24.3 28.7 43.1#
71 23.1 26.6 39.8#




Abundance Ion 42.00 (41.70 to 42.70): VW
Ion 72.00 (71.70 to 72.70): VW
Ion 71.00 (70.70 to 71.70): VW






#31
Cyclohexane
Concen: 1.254 ug/l
RT: 7.95 min Scan# 991
Delta R.T. -0.01 min
Lab File: VW010794.D
Acq: 13 Jun 2019 13:35

Tgt Ion: 56 Resp: 5739
Ion Ratio Lower Upper
56 100
69 145.7 25.2 37.8#
84 108.9 62.2 93.2#



Abundance Ion 56.00 (55.70 to 56.70): VW
Ion 69.00 (68.70 to 69.70): VW
Ion 84.00 (83.70 to 84.70): VW









#33
 1,2-Dichloroethane-d4
 Concen: 47.958 ug/l
 RT: 8.31 min Scan# 1050
 Delta R.T. -0.00 min
 Lab File: VW010794.D
 Acq: 13 Jun 2019 13:35




Instrument :
 MSVOA_W
 ClientSampled :


Tot Ion: 65 Resp: 108796
 Ion Ratio Lower Upper
 65 100
 67 51.6 0.0 102.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.84 min Scan# 1138
 Delta R.T. -0.00 min
 Lab File: VW010794.D
 Acq: 13 Jun 2019 13:35

Tgt Ion: 114 Resp: 327685
 Ion Ratio Lower Upper
 114 100
 63 24.6 0.0 46.4
 88 16.9 0.0 32.2





#35

Dibromofluoromethane

Concen: 0.866 ug/l

RT: 7.88 min Scan# 981

Delta R.T. -0.00 min

Lab File: VW010794.D

Acq: 13 Jun 2019 13:35

Instrument :
MSVOA_W
ClientSampleId :


Tot Ion:113 Resp: 1541

Ion Ratio Lower Upper




113 100

111 98.3 82.6 123.8

192 17.3 16.4 24.6



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 5.132 ug/l

RT: 7.95 min Scan# 991

Delta R.T. -0.13 min

Lab File: VW010794.D

Acq: 13 Jun 2019 13:35


Tgt Ion: 75 Resp: 17214

Ion Ratio Lower Upper



75 100


110 8.2 15.8 47.3#

77 0.0 24.7 37.1#



Abundance Ion 74.90 (74.60 to 75.60): VW
Ion 109.90 (109.60 to 110.60): V
Ion 76.90 (76.60 to 77.60): VW





#37

Ethyl Acetate

Concen: 2.925 ug/l

RT: 7.18 min Scan# 866

Delta R.T. -0.07 min

Lab File: VW010794.D

Acq: 13 Jun 2019 13:35

Instrument :

MSVOA_W

ClientSampleId :


Tot Ion: 43 Resp: 5437

Ion Ratio Lower Upper




43 100

61 0.0 9.7 14.5#

70 0.0 7.2 10.8#



Abundance Ion 43.00 (42.70 to 43.70): VW
Ion 60.90 (60.60 to 61.60): VW
Ion 70.00 (69.70 to 70.70): VW



#38

Carbon Tetrachloride

Concen: 6.697 ug/l

RT: 7.95 min Scan# 991

Delta R.T. -0.12 min

Lab File: VW010794.D

Acq: 13 Jun 2019 13:35


Tgt Ion: 117 Resp: 19392

Ion Ratio Lower Upper



117 100


119 3.5 74.2 111.4#

121 0.0 22.9 34.3#



Abundance Ion 116.80 (116.50 to 117.50): V
Ion 118.80 (118.50 to 119.50): V
Ion 120.80 (120.50 to 121.50): V









#48
Methvl methacrvlate
Concen: 1.698 ug/l
RT: 9.41 min Scan# 1232
Delta R.T. -0.02 min
Lab File: VW010794.D
Acq: 13 Jun 2019 13:35




Instrument :
MSVOA_W
ClientSampleId :


Tot Ion: 41 Resp: 2939
Ion Ratio Lower Upper
41 100
69 0.0 56.6 85.0#
39 42.0 44.4 66.6#



#50
Toluene-d8
Concen: 37.820 ug/l
RT: 10.32 min Scan# 1381
Delta R.T. -0.00 min
Lab File: VW010794.D
Acq: 13 Jun 2019 13:35

Tgt Ion: 98 Resp: 294974
Ion Ratio Lower Upper
98 100
100 65.4 53.0 79.6





#51

4-Methyl-2-Pentanone

Concen: 3.183 ug/l

RT: 10.21 min Scan# 1363

Delta R.T. 0.01 min

Lab File: VW010794.D

Acq: 13 Jun 2019 13:35

Instrument :

MSVOA_W





ClientSampled :

Tot Ion: 43 Resp: 5839

Ion Ratio Lower Upper

43 100

58 34.9 28.4 42.6



#59

2-Hexanone

Concen: 1.364 ug/l

RT: 10.98 min Scan# 1488

Delta R.T. 0.01 min

Lab File: VW010794.D




Acq: 13 Jun 2019 13:35


Tgt Ion: 43 Resp: 1808

Ion Ratio Lower Upper

43 100

58 68.7 24.9 74.6









#62
4-Bromofluorobenzene
Concen: 40.315 ug/l
RT: 12.62 min Scan# 1758
Delta R.T. -0.00 min
Lab File: VW010794.D
Acq: 13 Jun 2019 13:35




Instrument :
MSVOA_W
ClientSampleId :


Tot Ion: 95 Resp: 114418
Ion Ratio Lower Upper
95 100
174 74.4 0.0 150.4
176 71.8 0.0 144.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.63 min Scan# 1596
Delta R.T. -0.00 min
Lab File: VW010794.D
Acq: 13 Jun 2019 13:35

Tgt Ion: 117 Resp: 288214
Ion Ratio Lower Upper
117 100
82 60.9 47.0 70.4
119 32.8 25.1 37.7









#67
Ethyl Benzene
Concen: 1.076 ug/l
RT: 11.73 min Scan# 1612
Delta R.T. -0.00 min
Lab File: VW010794.D
Acq: 13 Jun 2019 13:35




Instrument :
MSVOA_W
ClientSampled :


Tot Ion: 91 Resp: 11717
Ion Ratio Lower Upper
91 100
106 32.2 24.2 36.4



#68
m/p-Xylenes
Concen: 1.722 ug/l
RT: 11.83 min Scan# 1629
Delta R.T. -0.01 min
Lab File: VW010794.D
Acq: 13 Jun 2019 13:35

Tgt Ion: 106 Resp: 6716
Ion Ratio Lower Upper
106 100
91 205.7 166.5 249.7









#69
o-Xylene
Concen: 2.409 ug/l
RT: 12.16 min Scan# 1683
Delta R.T. -0.01 min
Lab File: VW010794.D
Acq: 13 Jun 2019 13:35




Instrument :
MSVOA_W
ClientSampled :


Tot Ion:106 Resp: 8908
Ion Ratio Lower Upper
106 100
91 232.9 108.9 326.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.56 min Scan# 1912
Delta R.T. -0.00 min
Lab File: VW010794.D
Acq: 13 Jun 2019 13:35

Tgt Ion:152 Resp: 129436
Ion Ratio Lower Upper
152 100
115 63.1 29.8 89.3
150 157.2 0.0 345.0









#74
N-amvl acetate
Concen: 2.318 ug/l
RT: 12.24 min Scan# 1695
Delta R.T. -0.04 min
Lab File: VW010794.D
Acq: 13 Jun 2019 13:35




Instrument :
MSVOA_W
ClientSampleId :


Tot Ion: 43 Resp: 7286
Ion Ratio Lower Upper
43 100
70 0.0 29.2 43.8#
55 9.1 17.5 26.3#
61 0.0 17.4 26.2#



#76
1,2,3-Trichloropropane
Concen: 45.976 ug/l
RT: 12.62 min Scan# 1758
Delta R.T. -0.15 min
Lab File: VW010794.D
Acq: 13 Jun 2019 13:35

Tgt Ion: 75 Resp: 62493
Ion Ratio Lower Upper
75 100
77 1.3 0.0 0.0#









#80
 1,3,5-Trimethylbenzene
 Concen: 2.580 ug/l
 RT: 12.94 min Scan# 1811
 Delta R.T. -0.00 min
 Lab File: VW010794.D
 Acq: 13 Jun 2019 13:35




Instrument :
 MSVOA_W
 ClientSampleId :


Tot Ion:105 Resp: 20539
 Ion Ratio Lower Upper
 105 100
 120 46.8 24.3 72.8



#81
 trans-1,4-Dichloro-2-butene
 Concen: 105.668 ug/l
 RT: 12.62 min Scan# 1758
 Delta R.T. 0.11 min
 Lab File: VW010794.D
 Acq: 13 Jun 2019 13:35

Tgt Ion: 75 Resp: 62493
 Ion Ratio Lower Upper
 75 100
 53 0.0 90.1 135.1#
 89 0.0 34.6 52.0#





#84

1,2,4-Trimethylbenzene

Concen: 8.008 ug/l

RT: 13.26 min Scan# 1862

Delta R.T. -0.00 min

Lab File: VW010794.D

Acq: 13 Jun 2019 13:35

Instrument :

MSVOA_W


ClientSampleId :

Tot Ion:105 Resp: 64155

Ion Ratio Lower Upper




105 100

120 44.5 22.5 67.5



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#85

sec-Butylbenzene

Concen: 6.521 ug/l

RT: 13.26 min Scan# 1862

Delta R.T. -0.13 min

Lab File: VW010794.D


Acq: 13 Jun 2019 13:35

Tgt Ion:105 Resp: 64155

Ion Ratio Lower Upper



105 100


134 0.0 9.7 28.9#



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V





#90

Hexachloroethane

Concen: 2.196 ug/l

RT: 14.10 min Scan# 2001

Delta R.T. 0.01 min

Lab File: VW010794.D

Acq: 13 Jun 2019 13:35

Instrument :

MSVOA_W


ClientSampled :

Tot Ion:117 Resp: 3460

Ion Ratio Lower Upper




117 100

201 0.0 38.8 116.4#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 2.975 ug/l

RT: 14.46 min Scan# 2060

Delta R.T. -0.02 min

Lab File: VW010794.D

Acq: 13 Jun 2019 13:35


Tgt Ion: 75 Resp: 946

Ion Ratio Lower Upper

75 100

155 0.0 38.8 116.4#



157 0.0 48.4 145.0#




Abundance Ion 74.90 (74.60 to 75.60): VW

Ion 154.80 (154.50 to 155.50): V

Ion 156.80 (156.50 to 157.50): V





#95

Naphthalene

Concen: 3.453 ug/l

RT: 15.37 min Scan# 2209

Delta R.T. -0.00 min

Lab File: VW010794.D

Acq: 13 Jun 2019 13:35

Instrument :

MSVOA_W

ClientSampleId :


Tot Ion:128 Resp: 16210

Ion Ratio Lower Upper

128 100

127 13.9 10.6 15.8

129 11.5 8.6 13.0



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

