



Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061421\
Data File : VW019203.D
Acq On : 14 Jun 2021 09:00
Operator : SY/VA
Sample : BFB08
Misc : 5.00G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB108

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM060221SMA.M
Title : SFAM01.0
Last Update : Wed Jun 16 01:01:36 2021



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.3	38491	PASS
75	95	30	80	53.5	101301	PASS
95	95	100	100	100.0	189355	PASS
96	95	5	9	6.5	12353	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	74.6	141187	PASS
175	174	5	9	7.5	10625	PASS
176	174	95	105	97.7	137893	PASS
177	176	5	10	5.9	8138	PASS

m/z	Abundance
37.80	85.0
39.00	104.0
40.00	253.0
40.90	137.0
42.10	62.0
44.00	432.0
55.00	86.0
56.90	120.0
168.80	71.0

Instrument :
MSVOA_W
ClientSampleId :
BFB108

m/z	Abundance
35.10	64.0
36.00	2256.0
37.00	11151.0
38.00	9575.0
39.10	4346.0
40.00	328.0
41.00	164.0
42.90	134.0
44.00	1645.0
45.00	2136.0
46.00	271.0
47.00	2495.0
48.00	1429.0
49.00	9567.0
50.00	44056.0
51.00	13672.0
52.10	547.0
53.10	203.0
55.10	573.0
56.10	3908.0
57.00	6302.0
57.90	288.0
59.00	109.0
60.00	2078.0
61.00	11365.0
62.00	10892.0
63.00	8097.0
64.00	1049.0
65.00	1116.0
66.10	102.0
67.00	540.0
68.00	22232.0
69.00	22920.0
70.00	1769.0
70.90	105.0
72.00	1139.0
73.00	10803.0
74.00	37664.0
75.00	111608.0
76.00	9699.0
77.00	1586.0
77.90	851.0
78.90	7659.0
80.00	1895.0
80.90	8139.0
81.90	1742.0
82.80	227.0
85.90	181.0
87.00	9072.0
88.00	8260.0
90.90	904.0
92.00	6029.0
93.00	9563.0
94.00	25168.0
95.00	208320.0
96.00	13593.0
97.00	580.0
102.90	197.0
103.90	906.0
104.80	342.0
105.90	839.0
106.90	142.0
109.70	200.0
110.90	203.0
111.90	164.0
112.90	190.0
114.90	257.0
116.00	666.0
116.90	1346.0
118.00	598.0
118.90	1059.0
123.00	59.0
123.90	94.0
124.90	233.0
128.00	663.0
128.90	370.0
129.90	694.0
130.90	302.0
133.10	411.0
133.90	163.0
134.90	383.0
136.80	344.0
139.90	71.0
140.90	2153.0
142.00	264.0
142.90	2105.0
144.00	206.0
144.90	290.0
145.90	313.0
147.00	195.0
147.90	439.0
148.90	181.0

Instrument :
MSVOA_W
ClientSampleId :
BFB108

150.00	195.0
152.00	98.0
152.70	121.0
154.00	218.0
154.90	299.0
156.00	104.0
156.80	287.0
158.80	110.0
159.10	168.0
160.80	264.0
162.90	86.0
168.70	64.0
169.40	68.0
171.90	1372.0
173.90	135040.0
175.00	10758.0
175.90	132096.0
176.90	8378.0
178.00	365.0
178.90	118.0
190.90	421.0
191.90	77.0
193.00	669.0
193.90	108.0
205.00	112.0
207.10	341.0
208.20	147.0
249.10	372.0
250.10	65.0
251.10	202.0
265.10	235.0
278.70	52.0
280.20	168.0
281.10	3437.0
282.00	1049.0
282.90	561.0
283.90	172.0

Instrument :
MSVOA_W
ClientSampleId :
BFB108

m/z	Abundance
36.00	2043.0
37.00	11470.0
38.00	9297.0
39.00	3972.0
40.00	262.0
41.00	85.0
42.00	69.0
42.90	196.0
44.00	1555.0
45.00	2007.0
46.00	165.0
47.10	2285.0
48.00	1444.0
49.00	9203.0
50.00	42808.0
51.00	12881.0
52.00	650.0
53.00	180.0
54.90	696.0
56.00	3670.0
57.00	6995.0
58.00	322.0
59.10	59.0
60.00	2020.0
61.00	10788.0
62.00	10869.0
63.00	8987.0
64.00	967.0
65.10	918.0
65.90	96.0
67.00	571.0
68.00	22552.0
69.00	22392.0
70.00	1912.0
71.00	163.0
72.00	1055.0
73.00	9910.0
74.00	38536.0
75.00	113888.0
76.00	9346.0
76.90	1491.0
77.90	743.0
78.90	7279.0
79.90	1996.0
80.90	7094.0
81.90	1394.0
82.90	186.0
85.90	324.0
87.00	9728.0
88.00	9153.0
91.00	786.0
92.00	6402.0
93.00	9426.0
94.00	26264.0
95.00	210752.0
96.00	14243.0
96.90	392.0
102.80	86.0
103.90	859.0
104.90	345.0
105.80	962.0
106.90	213.0
110.00	99.0
111.00	248.0
111.90	189.0
113.00	148.0
115.00	249.0
115.80	799.0
116.90	1298.0
117.90	807.0
118.90	1105.0
123.00	63.0
123.90	105.0
124.70	122.0
125.90	115.0
127.80	658.0
128.90	372.0
129.90	873.0
130.80	276.0
133.00	237.0
134.00	135.0
135.00	333.0
136.70	299.0
139.80	74.0
140.90	1740.0
142.90	1847.0
144.00	143.0
144.90	295.0
145.70	254.0
147.00	177.0
147.80	448.0
148.90	132.0

Instrument :
MSVOA_W
ClientSampleId :
BFB108

149.90	203.0
151.80	64.0
152.60	113.0
153.90	155.0
155.00	592.0
156.10	112.0
157.00	327.0
158.00	135.0
159.00	182.0
160.90	151.0
161.80	68.0
170.80	153.0
171.10	138.0
172.00	1290.0
173.90	166656.0
175.00	12128.0
175.90	157632.0
177.00	9082.0
177.90	298.0
179.00	58.0
191.00	249.0
192.00	79.0
192.90	224.0
193.90	86.0
207.10	286.0
248.80	138.0
250.10	110.0
251.00	116.0
264.90	152.0
281.10	2721.0
282.10	646.0
283.10	460.0
284.00	137.0

Instrument :
MSVOA_W
ClientSampleId :
BFB108

m/z	Abundance
36.00	1271.0
37.00	6910.0
38.00	6436.0
39.10	2477.0
40.00	312.0
41.20	78.0
43.00	83.0
44.00	1423.0
45.00	1163.0
46.10	123.0
47.00	1792.0
47.90	883.0
49.00	6160.0
50.00	28608.0
51.00	8750.0
52.00	281.0
53.00	96.0
55.00	525.0
56.00	2566.0
57.00	4134.0
57.90	172.0
59.90	1802.0
61.00	7629.0
62.00	7468.0
63.00	5563.0
64.00	632.0
65.10	600.0
65.90	127.0
67.00	493.0
68.00	15848.0
69.00	15502.0
70.00	1371.0
70.90	133.0
71.90	861.0
73.00	7170.0
74.00	27008.0
75.00	78408.0
76.00	6487.0
76.90	941.0
78.00	788.0
78.90	5069.0
79.90	1571.0
80.90	5098.0
81.90	1186.0
83.00	132.0
86.00	269.0
87.00	6390.0
87.90	5316.0
89.20	80.0
90.90	488.0
92.00	4013.0
93.00	6682.0
94.00	17944.0
95.00	148992.0
96.00	9224.0
97.00	227.0
102.90	84.0
103.90	789.0
105.00	277.0
105.80	712.0
106.90	151.0
109.80	72.0
110.90	291.0
111.90	144.0
113.00	71.0
115.00	165.0
115.80	639.0
116.90	1129.0
117.80	452.0
118.90	840.0
122.90	63.0
123.80	70.0
124.90	114.0
125.90	82.0
127.90	556.0
129.00	306.0
129.90	600.0
130.80	195.0
132.90	126.0
134.90	248.0
136.90	250.0
140.00	93.0
141.00	1722.0
142.00	213.0
142.90	1719.0
144.00	78.0
144.90	197.0
145.80	347.0
146.70	67.0
147.00	57.0
147.80	258.0
148.70	132.0

Instrument :
MSVOA_W
ClientSampleId :
BFB108

149.90	122.0
151.80	95.0
152.90	175.0
153.80	133.0
154.90	469.0
156.00	110.0
157.00	218.0
158.90	164.0
160.70	209.0
171.10	240.0
171.90	870.0
173.90	121864.0
175.00	8988.0
175.90	123952.0
176.90	6954.0
178.00	134.0
191.00	77.0
193.00	60.0
194.00	51.0
206.80	88.0
207.10	84.0
248.80	83.0
265.00	161.0
281.10	1225.0
281.90	233.0
282.20	225.0
283.10	177.0

Instrument :
MSVOA_W
ClientSampleId :
BFB108