

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062018\
 Data File : VW003370.D
 Acq On : 20 Jun 2018 23:12
 Operator : SY/AP
 Sample : J3569-15
 Misc : 7.39G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DAXR7

Quant Time: Jun 21 05:06:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM062018S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 21 05:02:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	343710	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	320047	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	133607	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	115817	24.85	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	99.40%
7) Chloroethane-d5	2.89	69	89521	23.02	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.08%
10) 1,1-Dichloroethene-d2	4.01	63	161962	17.50	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	70.00%
20) 2-Butanone-d5	7.09	46	71745	48.84	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.68%
24) Chloroform-d	7.65	84	218448	24.16	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	96.64%
26) 1,2-Dichloroethane-d4	8.31	65	140240	26.10	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	104.40%
29) Benzene-d6	8.28	84	420625	24.77	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	99.08%
33) 1,2-Dichloropropane-d6	9.28	67	136558	25.40	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.60%
37) Toluene-d8	10.33	98	392521	24.18	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	96.72%
38) trans-1,3-Dichloropropene-	10.58	79	59305	23.50	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	94.00%
39) 2-Hexanone-d5	10.93	63	54853	49.93	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.86%
48) 1,1,2,2-Tetrachloroethane-	12.70	84	119137	23.82	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	95.28%
61) 1,2-Dichlorobenzene-d4	13.86	152	130327	25.83	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	103.32%

Target Compounds					Ovalue
13) Acetone	4.14	43	7459	4.76 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062018\
Data File : VW003370.D
Acq On : 20 Jun 2018 23:12
Operator : SY/AP
Sample : J3569-15
Misc : 7.39G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DAXR7

Quant Time: Jun 21 05:06:14 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM062018S.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 21 05:02:42 2018
Response via : Initial Calibration

