

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062019\
 Data File : VW010899.D
 Acq On : 20 Jun 2019 18:31
 Operator : SY/VA
 Sample : K3447-03
 Misc : 4.99G/5ML/MSVOA W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
ClientSampleId :
 SIDEWALL-1

Quant Time: Jun 21 08:37:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 05 03:21:19 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	238502	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	390651	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	330642	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	133545	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	148768	52.92	ug/l	0.00
Spiked Amount				50.000		
				Recovery	=	105.84%
35) Dibromofluoromethane	7.88	113	111678	52.67	ug/l	0.00
Spiked Amount				50.000		
				Recovery	=	105.34%
50) Toluene-d8	10.32	98	482686	51.91	ug/l	0.00
Spiked Amount				50.000		
				Recovery	=	103.82%
62) 4-Bromofluorobenzene	12.62	95	157655	46.60	ug/l	0.00
Spiked Amount				50.000		
				Recovery	=	93.20%
Target Compounds						
16) Acetone	4.14	43	6150	8.165	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062019\
 Data File : VW010899.D
 Acq On : 20 Jun 2019 18:31
 Operator : SY/VA
 Sample : K3447-03
 Misc : 4.99G/5ML/MSVOA W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SIDEWALL-1

Quant Time: Jun 21 08:37:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W060419S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 05 03:21:19 2019
 Response via : Initial Calibration

