

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW062119\  
 Data File : VW010915.D  
 Acq On : 22 Jun 2019 03:09  
 Operator : SY/VA  
 Sample : VW0621SBL01  
 Misc : 5.00G/10ML/MSVOA W/SOIL  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 VBLK23

Quant Time: Jun 22 06:46:06 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_W\METHOD\SOM2WLM062119S.M  
 Quant Title : VOC Analysis  
 QLast Update : Sat Jun 22 06:38:40 2019  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Difluorobenzene     | 8.84  | 114  | 356807   | 25.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 11.63 | 117  | 339468   | 25.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 13.56 | 152  | 159581   | 25.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |        |       |          |          |      |         |
|--------------------------------|--------|-------|----------|----------|------|---------|
| 4) Vinyl Chloride-d3           | 2.35   | 65    | 94987    | 21.79    | ug/L | 0.00    |
| Spiked Amount                  | 25.000 | Range | 30 - 150 | Recovery | =    | 87.16%  |
| 7) Chloroethane-d5             | 2.89   | 69    | 77031    | 21.47    | ug/L | 0.00    |
| Spiked Amount                  | 25.000 | Range | 30 - 150 | Recovery | =    | 85.88%  |
| 10) 1,1-Dichloroethene-d2      | 4.02   | 63    | 160796   | 15.98    | ug/L | 0.00    |
| Spiked Amount                  | 25.000 | Range | 45 - 110 | Recovery | =    | 63.92%  |
| 20) 2-Butanone-d5              | 7.07   | 46    | 124207   | 50.98    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 20 - 135 | Recovery | =    | 101.96% |
| 24) Chloroform-d               | 7.65   | 84    | 218323   | 21.94    | ug/L | 0.00    |
| Spiked Amount                  | 25.000 | Range | 40 - 150 | Recovery | =    | 87.76%  |
| 26) 1,2-Dichloroethane-d4      | 8.31   | 65    | 154592   | 23.87    | ug/L | 0.00    |
| Spiked Amount                  | 25.000 | Range | 70 - 130 | Recovery | =    | 95.48%  |
| 29) Benzene-d6                 | 8.27   | 84    | 458278   | 22.52    | ug/L | 0.00    |
| Spiked Amount                  | 25.000 | Range | 20 - 135 | Recovery | =    | 90.08%  |
| 33) 1,2-Dichloropropane-d6     | 9.27   | 67    | 156248   | 22.59    | ug/L | 0.00    |
| Spiked Amount                  | 25.000 | Range | 70 - 120 | Recovery | =    | 90.36%  |
| 37) Toluene-d8                 | 10.32  | 98    | 407044   | 22.03    | ug/L | 0.00    |
| Spiked Amount                  | 25.000 | Range | 30 - 130 | Recovery | =    | 88.12%  |
| 38) trans-1,3-Dichloropropene- | 10.58  | 79    | 62397    | 21.81    | ug/L | 0.00    |
| Spiked Amount                  | 25.000 | Range | 30 - 135 | Recovery | =    | 87.24%  |
| 39) 2-Hexanone-d5              | 10.93  | 63    | 77758    | 47.85    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 20 - 135 | Recovery | =    | 95.70%  |
| 48) 1,1,2,2-Tetrachloroethane- | 12.69  | 84    | 139523   | 23.64    | ug/L | 0.00    |
| Spiked Amount                  | 25.000 | Range | 45 - 120 | Recovery | =    | 94.56%  |
| 61) 1,2-Dichlorobenzene-d4     | 13.85  | 152   | 147006   | 23.16    | ug/L | 0.00    |
| Spiked Amount                  | 25.000 | Range | 75 - 120 | Recovery | =    | 92.64%  |

| Target Compounds | Ovalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW062119\  
Data File : VW010915.D  
Acq On : 22 Jun 2019 03:09  
Operator : SY/VA  
Sample : VW0621SBL01  
Misc : 5.00G/10ML/MSVOA W/SOIL  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_W  
ClientSampleId :  
VBLK23

Quant Time: Jun 22 06:46:06 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_W\METHOD\SOM2WLM062119S.M  
Quant Title : VOC Analysis  
QLast Update : Sat Jun 22 06:38:40 2019  
Response via : Initial Calibration

